

<same as the Log Summary field below>

IHE Change Proposal

Tracking information:

IHE Domain	Patient Care Device (PCD).
Change Proposal ID:	CP-PCD-054
Change Proposal Status:	Ballot
Date of last update:	12April2011
Person assigned:	John Rhoads

Change Proposal Summary information:

Require ORC-9 in all PCD-03 transactions	
Submitter's Name(s) and e-mail address(es):	Brad Lunde
Submission Date:	12April2011
Integration Profile(s) affected:	PIV
Actor(s) affected:	IOC and IOP.
IHE Technical Framework or Supplement modified:	Revision 1.2; September 30, 2010
Volume(s) and Section(s) affected:	Volume 2
Rationale for Change: To ensure messages being delivered from the IOP to the IOC are timely and old transactions can be identified easily and without interpretation it is requested that the ORC-9 be required for all PCD-03 transactions. The time in this field should be the time the clinician initiated the program request, not the time the IOP generated the message.	

Update Row 10 in table B.9-1:

SEQ	LEN	DT	Usage	Card.	TBL#	ITEM#	ELEMENT NAME
1	2	ID	R	[1..1]	0119	00215	Order Control
2	427	EI	C	[0..1]		00216	Placer Order Number
3	427	EI	R	[1..1]		00217	Filler Order Number
4	22	EI	RE	[0..1]		00218	Placer Group Number
5	2	ID	RE	[0..1]	0038	00219	Order Status
6	1	ID	RE	[0..1]	0121	00220	Response Flag
7	705	TQ	X	[0..0]		00221	Quantity/Timing
8	200	EIP	RE	[0..1]		00222	Parent
9	24	DTM	RE	[01..1]		00223	Date/Time of Transaction
10	3220	XCN	RE	[0..*]		00224	Entered By
11	250	XCN	RE	[0..*]		00225	Verified By
12	3220	XCN	RE	[0..*]		00226	Ordering Provider
13	80	PL	RE	[0..1]		00227	Enterer's Location
14	250	XTN	RE	[0..2]		00228	Call Back Phone Number
15	24	DTM	RE	[0..1]		00229	Order Effective Date/Time
16	250	CE	RE	[0..1]		00230	Order Control Code Reason
17	250	CE	RE	[0..1]		00231	Entering Organization
18	250	CE	RE	[0..1]		00232	Entering Device
19	250	XCN	R	[1..1]		00233	Action By

<same as the Log Summary field below>

SEQ	LEN	DT	Usage	Card.	TBL#	ITEM#	ELEMENT NAME
20	250	CE	RE	[0..1]	0339	01310	Advanced Beneficiary Notice Code
21	250	XON	RE	[0..*]		01311	Ordering Facility Name
22	250	XAD	RE	[0..*]		01312	Ordering Facility Address
23	250	XTN	RE	[0..*]		01313	Ordering Facility Phone Number
24	250	XAD	RE	[0..*]		01314	Ordering Provider Address
25	250	CWE	RE	[0..1]		01473	Order Status Modifier
26	60	CWE	RE	[0..1]	0552	01641	Advanced Beneficiary Notice Override Reason
27	24	DTM	RE	[0..1]		01642	Filler's Expected Availability Date/Time
28	250	CWE	RE	[0..1]	0177	00615	Confidentiality Code
29	250	CWE	RE	[0..1]	0482	01643	Order Type
30	250	CNE	RE	[0..1]	0483	01644	Enterer Authorization Mode

Replace Section B.9 lines 2177-2179 :

ORC-9 Date/Time of Transaction

See HL7 V2.6 Section 4.5.1.9 for details. The PCD TF does not further constrain this field.

ORC-9 Date/Time of Transaction

The time in this field should be the time the clinician initiated the program request, not the time the IOP generated the message. The IOC may use this field to determine if the request is stale or too old.

Add a row to the PIV error code table lines(??); 90XX should be the next logical error code in the table:

90XX	Requested infusion program is stale.	The value of ORC.9 is older than what the IOC will allow to program the pump.
-------------	--------------------------------------	---