Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-PCD-076-00
Change Proposal Status:	Ballot
Date of last update:	2012-06-05
Person assigned:	John Rhoads

Change Proposal Summary information:

Numbering OBRs	
Submitter's Name(s) and e-mail address(es):	
Submission Date:	2011-11-03
Integration Profile(s) affected:	General
Actor(s) affected:	General
IHE Technical Framework or Supplement modified:	Final Text, August 11, 2011
Volume(s) and Section(s) affected:	V. 2, App. B Sec. 7
Rationale for Change: Text for OBR-1 Set ID OBR did not explicitly state that the numbering of OBR segments in that field restarts at 1 in each new message	

Insert new text shown in bold underline.

OBR-1   Set ID OBR

Definition: For the first order transmitted **in each message**, the sequence number shall be 1; for the second order, it shall be 2; and so on.