

&lt;same as the Log Summary field below&gt;

# IHE Change Proposal

## Tracking information:

IHE Domain	Patient Care Device
Change Proposal ID:	CP-PCD-075-00
Change Proposal Status:	Submitted
Date of last update:	2012.05.25
Person assigned:	John Rhoads

## Change Proposal Summary information:

Add RTM Specification Detail to the PCD TF-3	
	Todd Cooper (ToddCooperAFC@gmail.com)
Submission Date:	2012.05.25
Integration Profile(s) affected:	RTM Content Module
Actor(s) affected:	None
IHE Technical Framework or Supplement modified:	PCD
Volume(s) and Section(s) affected:	Primarily TF-3, section 4 (which is reserved for RTM)
<p><b>Rationale for Change:</b></p> <p>Though the Rosetta Terminology Mapping (RTM) constrained data set is mandatory for all PCD transaction content, the detailed specification information is not currently included in the PCD TF. There is mention of it in the following sections:</p> <p style="padding-left: 40px;">TF-1 (2.2.4) – This high-level overview is misplaced, since RTM is not strictly a Profile; a new content module section should be created</p> <p style="padding-left: 40px;">TF-1 (Appendix A) – RTM Problem Statement &amp; Key Use Case; this content should be moved to TF-3 (4.1)</p> <p style="padding-left: 40px;">TF-2 (Appendix K) – Additional overview and background information about RTM, some of which is now dated with the public release of the NIST RTMMS tool.</p> <p>None of these sections include the detailed specifications that define the RTM design. As noted, the information is also both dated and in some cases, not in the appropriate location.</p> <p>This CP updates the sections identified above as well as adds the detailed information to TF-3 (4.1) to capture RTM detailed specifications.</p>	