

# IHE Change Proposal

## Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-PCD-065
Change Proposal Status:	Submitted
Date of last update:	2012-05-22
Person assigned:	John Rhoads

## Change Proposal Summary information:

Editorial Update - Glossary	
Submitter's Name(s) and e-mail address(es):	Stephen Swanson ( <a href="mailto:sswanson@sjm.com">sswanson@sjm.com</a> ) Gene Hageman ( <a href="mailto:gene.hageman@medtronic.com">gene.hageman@medtronic.com</a> ) Alexander Kraus ( <a href="mailto:alexander.kraus@biotronik.com">alexander.kraus@biotronik.com</a> ) Ken Hoyme ( <a href="mailto:ken.hoyme@bsci.com">ken.hoyme@bsci.com</a> )
Submission Date:	April 05, 2012
Integration Profile(s) affected:	Implantable Device – Cardiac – Observation
Actor(s) affected:	None
IHE Technical Framework or Supplement modified:	IHE Patient Care Device (PCD) Technical Framework, Volume 1; Revision 1.0; August 12, 2011
Volume(s) and Section(s) affected:	Volume 1 (PCD TF-1); Glossary
Rationale for Change: 1. Modify definition of 'Cardiac Device Programmer' to be specific for implantable devices. 2. Add one new definition for a 'Implantable Cardiac Monitor' which is supported by the IDCO profile.	

## Proposed Changes:

*Modify wording in Glossary with the following:*

- Cardiac Device Programmer – A device used to noninvasively interrogate, monitor, and alter the operating parameters of an implantable pacemaker, implantable defibrillator, ~~or~~ implantable cardiac resynchronization device, or implantable cardiac monitor.

*Add to the Glossary the following:*

- Implantable Cardiac Monitor – An electronic device implanted beneath the skin for monitoring/recording the electrical signals of the heart muscle.