

IHE Change Proposal

Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-PCD-059
Change Proposal Status:	Ballot
Date of last update:	2011-06-13
Person assigned:	John Rhoads

Change Proposal Summary information:

Technical Framework Final Text Editorial Changes	
Submitter's Name(s) and e-mail address(es):	John Rhoads (john.rhoads@philips.com)
Submission Date:	2011-06-13
Integration Profile(s) affected:	PIV, DEC
Actor(s) affected:	
IHE Technical Framework or Supplement modified:	IHE Patient Care Device Technical Framework Volume 2 Transactions
Volume(s) and Section(s) affected:	See below
Rationale for Change: PIV HL7 Enhanced Acknowledgement Mode passed 2011 Connectathon and met checklist criteria by Planning Committee vote, so integrated into Final Text version. By request of some Technical Committee members, discussion of OBX-4 Sub-id and supporting information edited for clarity.	

Specific proposed changes to text relative to version of 2011-04-01, which was previously balloted successfully, may be seen by viewing file IHE_PCD_TF_Vol2_FT_2011-06-13TrackChanges.doc with Microsoft Word Track Changes tool active.

2.3	made fuller reference to Rosetta Terminology Mapping, Appendix K
3.1.4.1.1	removed mention of subscriptions – SPD not yet in Final Text
3.1.4.2.2	Trigger Events. Clarified that DOR "shall not" do interpolation.
3.1.4.4	Correct typo – "validate" -> "validates"
3.3.4.2	Correct typo –PV1 is permitted, implying cardinality is [0..1]
3.3.4.1.1	Added HL7 Enhanced Acknowledgement material from previously approved PIV Trial Implementation Supplement
3.3.4.4.5	Corrected typo – Sub-id is required, so cardinality is [1..1]
3.3.4.4.2	Corrected type – Administrative device is "R", not "RE"
3.3.4.4.9	Material added from PIV TI Supplement on HL7 Enhanced Acknowledgment Mode
App. A	Edited to be consistent with OBX-4 discussion in Appendix B

<same as the Log Summary field below>

App. A.2	Material out of place – move to App. B
App. B.1	Editorial clarification
App. B.2	Material added from PIV TI Supplement on HL7 Enhanced Acknowledgment Mode
App. B.4.1	Typo – Administrative device - wrong 2.6 field length. Corrected to 705
App. B	Editorial clarification on OBR-4, OBX-4, Time stamps
App. K	New appendix with material moved in to "centralize" discussion of Rosetta Terminology Management