

IHE Change Proposal

Tracking Information

IHE Domain	Patient Care Devices
Change Proposal ID	CP-PCD-040-00
Change Proposal Status	Ballot
Date of last update	2010-12-21
Person assigned	J. Rhoads

Change Proposal Summary Information

Title:	Put in proper increased HL7 v2.6 OBX field widths in Appendix B
Submitter's Name(s) and e-mail address(es):	S. Eaton
Submission Date:	2010-12-21
Integration Profile(s) affected:	DCE, PIV
Actor(s) affected:	DOR, DOC, IOR, IOC
IHE Technical Framework or Supplement modified:	Technical Framework Trial Implementation
Volume(s) and Section(s) affected:	vol. 2, Table 3.3.4.4.8-1, Table B.8.2
Rationale for Change: Previous Trial Implementation Technical Framework was based on HL7 2.5. New version uses 2.6. Numerous field length limits were increased in v. 2.6.	

Not all of these changes were reflected properly in table B.8.2 in section B.8. This change introduces the corrected lengths for OBX-2, OBX-15, OBX-16, OBX-17 and OBX-18. In table 3.3.4.4.8-1, length of data type field OBX-2.