

IHE Change Proposal

Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-PCD-068
Change Proposal Status:	Submitted
Date of last update:	2012-05-22
Person assigned:	John Rhoads

Change Proposal Summary information:

PID 3.1 Description not updated	
Submitter's Name(s) and e-mail address(es):	Stephen Swanson (sswanson@sjm.com) Gene Hageman (gene.hageman@medtronic.com) Alexander Kraus (alexander.kraus@biotronik.com) Ken Hoyme (ken.hoyme@bsci.com)
Submission Date:	April 05, 2012
Integration Profile(s) affected:	Implantable Device – Cardiac – Observation
Actor(s) affected:	Implantable Device – Cardiac – Reporter Implantable Device – Cardiac – Consumer
IHE Technical Framework or Supplement modified:	IHE Patient Care Device (PCD) Technical Framework; Volume 2; Revision 1.0; August 12, 2011
Volume(s) and Section(s) affected:	Volume 2 (PCD TF-2); Section 3.9.4.1.2.7
Rationale for Change: 1. Updates were applied to the IEEE nomenclature which were not carried over to the IHE profile where the data type for the 'Data' parameter. This changed the parameter from TX to ED.	

Proposed Changes:

Replace text in Table 3.9.4.1.2.7-1 Row 11 by the following:

Table 3.9.4.1.2.7-1: OBX Segment

Name	Seq	DT	Len	Opt	Rep	Min	Max	Tbl	Fixed Value	Ex Val
Data	5	TX ED	*	RE		1	1		Y	Encapsulated and Base64 binary encoded PDF File