

IHE Change Proposal

Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-PCD-070-00
Change Proposal Status:	Submitted
Date of last update:	8-12-2011
Person assigned:	John Rhoads

Change Proposal Summary information:

Revised OBX-5 Explanation	
Submitter's Name(s) and e-mail address(es):	Stephen Swanson (sswanson@sjm.com) Gene Hageman (gene.hageman@medtronic.com) Alexander Kraus (alexander.kraus@biotronik.com) Ken Hoyme (ken.hoyme@bsci.com)
Submission Date:	April 05, 2012
Integration Profile(s) affected:	Implantable Device – Cardiac – Observation
Actor(s) affected:	None
IHE Technical Framework or Supplement modified:	IHE Patient Care Device (PCD) Technical Framework; Volume 2; Revision 1.0; August 12, 2011
Volume(s) and Section(s) affected:	Volume 2 (PCD TF-2); Section 3.9.4.1.2.4
Rationale for Change:	
1. The example value provided in Table 3.2.4.1.4-2: OBR Segment does not match the value defined in the IEEE 11073 Nomenclature. The Change Proposal is to bring the example text in-line with the IEEE value.	

Proposed Changes:

Replace Table 3.2.4.1.4-2 Row 7 by the following:

Table 3.2.4.1.4-2: OBR Segment

Name	Seq	DT	Len	Opt	Rep	Min	Max	Tbl	Fixed Val	Ex Val
<i>identifier</i>	1	ST	20	R		1	1			Remote follow-up

IHE Change Proposal

Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-DOM-xxx (assigned by Domain Technical Committee)
Change Proposal Status:	Submitted
Date of last update:	8-12-2011
Person assigned:	(assigned by Domain Technical Committee)

Change Proposal Summary information:

Example Text Change	
Submitter's Name(s) and e-mail address(es):	Stephen Swanson (sswanson@sjm.com) Gene Hageman (gene.hageman@medtronic.com) Alexander Kraus (alexander.kraus@biotronik.com) Ken Hoyme (ken.hoyme@bsci.com)
Submission Date:	May 23, 2012
Integration Profile(s) affected:	Implantable Device – Cardiac – Observation
Actor(s) affected:	None
IHE Technical Framework or Supplement modified:	IHE Patient Care Device (PCD) Technical Framework; Volume 2; Revision 1.0; August 12, 2011
Volume(s) and Section(s) affected:	Volume 2 (PCD TF-2); Section 3.9.4.1.2.5
Rationale for Change: 1. There was confusion on what values should be contained in the OBX-5 fields when the Value Type (OBX-2) is Coded With Exceptions (CWE). The Change Proposal clarifies the intended use of this field.	

Proposed Changes:

Replace the text defining the use of OBX-5 with the following:

OBX-5 Observation Value – This is the actual value of the observation.

If OBX-2 is of type CWE then

OBX-5.1 shall be an *enumeration code [mnemonic]* as defined in Annex D.3 ‘enumerations’ or Annex E.3 ‘vendor enumerations’ of IEEE 11073-10103 version 1.05.02 (10/04/2011) and OBX-5.2 can contain the according *display name* as defined in Annex D.3 ‘enumerations’ or Annex E.3 ‘vendor enumerations’ of IEEE 11073-10103 version 1.05.02 (10/04/2011) or an equivalent display name in another language.