

### Tracking information:

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-PCD-051
Change Proposal Status:	Ballot
Date of last update:	2010-11-30
Person assigned:	Ken Fuchs

### Change Proposal Summary information:

Changes to IHE_PCD_TF_Vol1_FT_2010-10-18Draft	
Submitter's Name(s) and e-mail address(es):	Ken Fuchs
Submission Date:	2010-11-30
Integration Profile(s) affected:	IHE PCD TF Volume 1
Actor(s) affected:	Various
IHE Technical Framework or Supplement modified:	IHE_PCD_TF_Vol1_FT_2010-10-18Draft
Volume(s) and Section(s) affected:	Various
<p>Rationale for Change:</p> <p>The IHE PCD TF Volume 1 Draft was published on 2010-10-20 for voting. As a result a number of comments were received and incorporated into this single change proposal.</p> <p>Even though the changes proposed are editorial in nature, and the TF FT did pass the voting process, it was decided to create this CP and resubmit an updated FT for another round of voting.</p>	

Section 1.7 – Improve description of RTM profile

Section 2.1 – Update wording to improve consistency with other Domains

Section 3.1 – Time Client is not an IHE-PCD Actor, remove from Existing Actors and remove from Table 3.1-1

Section 3.2 – Maintain Time is not an IHE-PCD Transaction, remove from list of Transactions and from Table 3.2-1

Section 4.1 – Fix reference to Section 4.2, and remove Maintain Time from Table 4.1-1

Section 4.2 – Remove Subscribe to PCD Data, and fix reference to MLLP Transport in Table 4.2-1

Section 5.1 – Add reference to PCD-03 and remove Maintain Time from Table 5.1-1

Section 6.2 – Add description of options and their purpose. Reformat Table 6.2-1 to improve consistency with other similar tables in the TF.

Glossary – Add definition of MLLP