IHE Change Proposal

Tracking Information

IHE Domain	Patient Care Devices
Change Proposal ID	CP-PCD-043
Change Proposal Status	Ballot
Date of last update	2010-12-21
Person assigned	John Rhoads

Change Proposal Summary Information

Title:	**Cardinality of ERR segment in RRG^O16**
Submitter's Name(s) and e-mail address(es):	B. Lunde
Submission Date:	2010-12-21
Integration Profile(s) affected:	PIV
Actor(s) affected:	IOC
IHE Technical Framework or Supplement modified:	Technical Framework Final Text
Volume(s) and Section(s) affected:	Vol. 2, Section 3.3.5
Rationale for Change: Correct a typographical error in moving the material from PIV Supplement to Technical Framework.	

Section 3.3.5. Change "R" to "C" and cardinality from [0:0] to [0:1] in synopsis of segments in message.