

<same as the Log Summary field below>

**Tracking information:**

IHE Domain	Patient Care Devices
Change Proposal ID:	CP-PCD-074-00
Change Proposal Status:	Ballot
Date of last update:	2012-06-06
Person assigned:	John Rhoads

**Change Proposal Summary information:**

Error in MSH-21 description	
Submitter's Name(s) and e-mail address(es):	
Submission Date:	2012-11-03
Integration Profile(s) affected:	General
Actor(s) affected:	General
IHE Technical Framework or Supplement modified:	
Volume(s) and Section(s) affected:	
Rationale for Change: Example inconsistent with recommendation in text.	

In the section on OIDs (around line 1270), it is recommended that EI-1 be in the form "IHE\_PCD\_XXX", however the example is "PCD\_DEC\_001".

Change line 1279 from:

Example: PCD\_DEC\_001^IHE PCD^1.3.6.1.4.1.19376.1.6.1.1.1^ISO

To:

Example: IHE\_PCD\_001^IHE PCD^1.3.6.1.4.1.19376.1.6.1.1.1^ISO