

*Indian Standard*

**METHODS FOR  
MEASUREMENT OF AIR POLLUTION  
PART XIX CHLORINE**

Third Reprint May 1995

UDC 614.71:628.512:543.272.4

*Copyright 1982*

**BUREAU OF INDIAN STANDARDS  
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG  
NEW DELHI 110002**

# Indian Standard

## METHODS FOR MEASUREMENT OF AIR POLLUTION

### PART XIX CHLORINE

Air Quality Sectional Committee, CDC 53

*Chairman*

DR B. B. SUNDARESAN

*Representing*

National Environmental Engineering Research  
Institute ( CSIR ), Nagpur

*Members*

SHRI P. K. YENNAWAR ( *Alternate to*  
Dr B. B. Sundaresan )

SHRI A. L. AGGARWAL

National Institute of Occupational Health ( ICMR ),  
Ahmadabad

DR J. S. AHLUWALIA

Indian Oil Corporation Ltd, Faridabad

SHRI V. S. MORE ( *Alternate* )

SHRI K. D. AMRE

National Organic Chemical Industries Ltd,  
Bombay

SHRI H. M. SHAIKH ( *Alternate* )

SHRI N. G. ASHAR

The Dharamsi Morarji Chemical Co Ltd, Bombay

DR M. S. VAIDYA ( *Alternate* )

SHRI A. K. BASU

Calcutta Metropolitan Development Authority  
Calcutta

SHRI RANJIT KUMAR  
SENGUPTA ( *Alternate* )

DR NILAY CHAUDHURI

Central Board for the Prevention and Control of  
Water Pollution, New Delhi

MEMBER SECRETARY ( *Alternate* )

SHRI DALJIT SINGH

Hindustan Steel Ltd, Ranchi

DR R. K. DUTTA ( *Alternate* )

SHRI J. M. DAVE

Jawaharlal Nehru University, New Delhi

DR P. J. DEORAS

Society for Clean Environment, Bombay

DR S. B. CHAPHEKAR ( *Alternate I* )

SHRI T. N. MAHADEVAN ( *Alternate II* )

SHRI M. V. DESAI

Indian Chemical Manufacturers Association,  
Calcutta

SHRI B. SARAN ( *Alternate* )

SHRI B. D. DESHMUKH

Maharashtra Prevention of Water Pollution Board,  
Bombay

SHRI QAISAR AZEEZ ( *Alternate* )

SHRI N. B. ENGINEER

Cement Manufacturers' Association, Bombay

( Continued on page 2 )

© Copyright 1982

BUREAU OF INDIAN STANDARDS

This publication is protected under the *Indian Copyright Act* ( XIV of 1957 ) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

( Continued from page 1 )

| <i>Members</i>                               | <i>Representing</i>  |
|--|--|
| SHRI A. K. GUPTA                             | Indian Institute of Petroleum ( CSIR ), Dehra Dun                            |
| SHRI B. P. PUNDIR ( <i>Alternate</i> )       |  |
| DR V. S. GUPTA                               | Fertilizer ( Planning & Development ) India Ltd,<br>Sindri                   |
| SMT M. CHANDRA ( <i>Alternate</i> )          |  |
| DR H. B. MATHUR                              | Indian Institute of Technology, New Delhi                                    |
| DR N. K. MEHROTRA                            | Industrial Toxicology Research Centre ( CSIR ),<br>Lucknow                   |
| DR P. N. VISHWANATHAN ( <i>Alternate I</i> ) |  |
| DR J. L. KAW ( <i>Alternate II</i> )         |  |
| SHRI R. S. MEHTA                             | Gujarat Water Pollution Control Board,<br>Gandhinagar                        |
| SHRI G. B. SONI ( <i>Alternate</i> )         |  |
| SHRI A. MOOKHERJEE                           | S. F. India Ltd, Calcutta  |
| SHRI M. CHAUDHURY ( <i>Alternate</i> )       |  |
| DR P. N. MUKHERJEE                           | Central Fuel Research Institute ( CSIR ), Dhanbad                            |
| DR R. U. ROY ( <i>Alternate</i> )            |  |
| DR V. PACHAIYAPAN                            | The Fertilizer Association of India, New Delhi                               |
| DR B. PADMANABHAMURTHY                       | Meteorological Department, New Delhi   |
| SHRI A. R. PANICKER                          | Hindustan Insecticides Ltd, New Delhi  |
| SHRI Y. G. PATANKAR                          | Union Carbide India Ltd, Calcutta  |
| SHRI J. D. PATEL ( <i>Alternate</i> )        |  |
| SHRI J. R. PATWARDHAN                        | Municipal Corporation of Greater Bombay                                      |
| SHRI V. B. SHIRODKAR ( <i>Alternate</i> )    |  |
| DR P. K. RAMACHANDRAN                        | Ministry of Defence, Delhi   |
| DR B. V. RAMANI ( <i>Alternate</i> )         |  |
| SHRI V. RAMADURAI                            | Alkali Manufacturers' Association of India,<br>Bombay                        |
| SHRI R. V. RAMANI ( <i>Alternate</i> )       |  |
| DR S. S. RAMASWAMY                           | Directorate General, Factory Advice Service and<br>Labour Institutes, Bombay |
| SHRI S. K. DANGWAL ( <i>Alternate</i> )      |  |
| SHRI A. N. RAO                               | Directorate General of Technical Development,<br>New Delhi                   |
| SHRI S. B. SARKAR                            | Coal India Ltd, Calcutta   |
| DR V. V. SHIRVAIKAR                          | Bhabha Atomic Research Centre, Bombay  |
| DR R. K. KAPOOR ( <i>Alternate</i> )         |  |
| DR J. K. SINHA                               | Central Mining Research Station ( CSIR ), Dhanbad                            |
| SHRI A. K. BOSE ( <i>Alternate</i> )         |  |
| SHRI S. A. SUBRAMANIAN                       | Central Electricity Authority, New Delhi                                     |
| DR HARI BHAGWAN,<br>Director ( Chem )        | Director General, ISI ( <i>Ex-officio Member</i> )                           |

*Secretary*

SHRI A. K. BAHL  
Assistant Director ( Chem ), ISI

( Continued on page 8 )

# *Indian Standard*

## METHODS FOR MEASUREMENT OF AIR POLLUTION

### PART XIX CHLORINE

#### 0. FOREWORD

**0.1** This Indian Standard ( Part XIX ) was adopted by the Indian Standards Institution on 20 May 1982, after the draft finalized by the Air Quality Sectional Committee had been approved by the Chemical Division Council.

**0.2** Chlorine gas is primarily a respiratory irritant. It is so irritating that concentrations above 9 mg/m<sup>3</sup> in air are readily detectable by a normal person

**0.3** Devices and techniques for determining the concentration of pollutants in the atmosphere are important for the assessment of ' ambient air quality ', establishing hazardous levels in the environment, ameliorative measures and appraisal of contamination from a process or source.

**0.4** In the preparation of this standard, considerable assistance has been derived from the Publication No. 42215-01-70 ' Tentative methods of analysis for free chlorine content of the atmosphere ( methyl orange method ) ' issued by the American Public Health Association.

**0.5** In reporting the result of a test or analysis made in accordance with this standard, if the final value, observed or calculated, is to be rounded off, it shall be done in accordance with IS : 2-1960\*.

---

#### 1. SCOPE

**1.1** This standard ( Part XIX ) describes the details of methyl orange method for measurement of free chlorine in air.

---

\*Rules for rounding off numerical values ( revised ).

## 2. METHYL ORANGE METHOD

**2.1 Principle** — The colour of methyl orange solution ceases to vary with acidity near a pH value of 3.0. The dye is quantitatively bleached by free chlorine, and the extent of bleaching may be determined colorimetrically. The optimum concentration range is 0.05-1.0 ppm in ambient air ( 145  $\mu\text{g}$  to 2 500  $\mu\text{g}$  per  $\text{m}^3$  at 25°C and 100 kPa ).

NOTE — It is essential to maintain proper pH

**2.2 Range and Sensitivity** — The procedure given is designed to cover the range of 5-100  $\mu\text{g}$  of free chlorine per 100 ml of sampling solution. For a 30 litre air sample, this corresponds to approximately 0.05-1.0 ppm in air which is the optimum range. Increasing the volume of air samples extends the range at the lower end, but only within limits, since 50 litre of chlorine free air produces the same effect as about 0.01 ppm of chlorine.

**2.3 Limitations and Interferences** — Free bromine, which gives the same reaction, interferes in a positive direction. Manganese ( III, IV ) in concentrations of 0.1 ppm or above also interferes positively. Negative interference from sulphur dioxide is significant both in solution as well as gaseous state. Nitrites impart an off-colour orange to the methyl orange reagent. Nitrogen dioxide interferes positively, reacting as 20 percent chlorine. Sulphur dioxide interferes negatively, decreasing the chlorine by an amount equal to one third the sulphur dioxide concentration.

**2.4 Precision and Accuracy** — Error due to measurement by this procedure is known to be less than  $\pm 5$  percent of the amount present.

### 2.5 Apparatus

**2.5.1 Spectrophotometer** — suitable for measurement at 505 nm, preferably accommodating 5-cm cells.

**2.5.2 Sampling Train** — according to IS : 5182 ( Part V )-1975\* using a large impinger with fritted disc [ see Fig. 2 of IS : 5182 ( Part V )-1975\* ] of porosity B ( 70 to 100  $\mu\text{m}$  maximum pore diameter ).

**2.6 Reagents** — Pure chemicals of analytical grade, unless otherwise specified, and distilled water ( see IS : 1070-1977† ) shall be employed in the test. These shall not contain any impurities that affect the results of analysis.

**2.6.1 Methyl Orange Stock Solution 0.05 Percent** — Dissolve 0.500 g of methyl orange in water and dilute to one litre. This solution is stable indefinitely if freshly boiled and cooled distilled water is used.

\*Methods for measurement of air pollution : Part V Sampling of gaseous pollutants.

†Specification for water for general laboratory use ( second revision ).

**2.6.2 Methyl Orange Reagent 0.005 Percent** — Dilute 100 ml of stock solution to one litre with water. Prepare fresh for use.

**2.6.3 Sampling Solution** — Dilute 6 ml of 0.005 percent methyl orange reagent to 100 ml with water and add three drops ( 0.15-0.20 ml ) of 5 N hydrochloric acid. One drop of butanol may be added to induce foaming and increase collection efficiency. pH of the solution may be checked using a pH meter.

**2.6.4 Acidified Water** — Add three drops ( 0.15-0.20 ml ) of 5 N hydrochloric acid to 100 ml of water.

**2.6.5 Potassium Dichromate Solution 0.100 0 N** — Dissolve 4.904 g anhydrous potassium dichromate primary standard grade in water and dilute to one litre.

**2.6.6 Starch Indicator Solution** — Prepare a thin paste of 1 g of soluble starch in a few millilitres of water. Bring 200 ml of water to boil, remove from heat, and stir in the starch paste. Prepare fresh before use.

**2.6.7 Potassium Iodide** — reagent grade.

**2.6.8 Sodium Thiosulphate Solution 0.1N** — Dissolve 25 g of  $\text{Na}_2\text{S}_2\text{O}_3 \cdot 5\text{H}_2\text{O}$  in freshly boiled and cooled water and dilute to one litre. Add 5 ml chloroform as preservative and allow to age for two weeks before standardizing. Take 80 ml water, 1 ml sulphuric acid and 1 g of potassium iodide in a titration flask and pipette into this, with constant stirring, 10 ml of 0.100 0 N potassium dichromate solution. Allow to stand in the dark for 6 minutes. Titrate with 0.1 N thiosulphate solution. Upon approaching the end-point ( brown colour changing to yellowish green ), add 1 ml of starch indicator solution and continue titrating to the end-point ( blue to light green ).

$$\text{Normality of } \text{Na}_2\text{S}_2\text{O}_3 = \frac{1.000\ 0}{V}$$

where

$V$  = volume in ml of sodium thiosulphate solution used.

**2.6.9 Sodium Thiosulphate Solution 0.01 N** — Dilute 100 ml of the aged and standardized 0.1 N sodium thiosulphate solution to 1 litre with freshly boiled and cooled distilled water. Add 5 ml chloroform as preservative and store in a glass-stoppered bottle. Standardize frequently with 0.010 0 N potassium dichromate.

**2.6.10 Chlorine Solution, Approximately 10 ppm** — Prepare by serial dilution of household bleach ( approximately 50 000 ppm ), or by dilution of strong chlorine water made by the bubbling chlorine gas through cold

distilled water. The diluted solution should contain approximately 10 ppm of free ( available ) chlorine. Prepare one litre.

**2.7 Procedure** — Assemble the sampling train. Add 100 ml of sampling solution to the impinger and draw a measured volume of air at a rate of 1-2 l/min for a period of time appropriate to the estimated chlorine concentration. Transfer the solution to a 100-ml volumetric flask and make to volume, if necessary, with acidified water. Measure absorbance at 505 nm in 5-cm cells against water as reference.

**2.7.1** The volume of sampling solution, the concentration of methyl orange in the sampling solution, the amount of air sampled, the size of impinger and the length of photometer cell may be varied to suit the needs of the situation as long as proper attention is paid to the corresponding changes necessary in the calibration procedure.

## 2.8 Calibration

**2.8.1** Prepare a series of six 100-ml volumetric flasks containing 6 ml of 0.005 percent methyl orange reagent, 75 ml water, and 3 drops ( 0.15-0.20 ml ) of 5.0 N hydrochloric acid. Carefully and accurately pipette 0, 0.5, 1.0, 5.0 and 9.0 ml of chlorine solution ( approximately 10 ppm ) into the respective flasks; holding the pipette tip beneath the surface. Quickly mix and make to volume with acidified water.

**2.8.2** Immediately standardize the 10 ppm chlorine solution by adding 400 ml of chlorine solution to a flask containing 1 g potassium iodide and 5 ml glacial acetic acid, and swirling to mix. Titrate with 0.01 N sodium thiosulphate until the iodine colour becomes a faint yellow. Add 1 ml of starch indicator solution and continue the titration to the end-point ( blue to colourless ). One millilitre of 0.010 0 N sodium thiosulphate is equivalent to 0.354 6 mg of free chlorine. Calculate the amounts of free chlorine added to each flask in **2.8.1**.

**2.8.3** Transfer the standards prepared in **2.8.1** to absorption cells and measure absorbance against micrograms of chlorine to draw the standard curve.

**2.9 Calculations** — Calculate the chlorine concentration as follows:

$$\text{Chlorine, } \mu\text{g/m}^3 = \frac{M}{V}$$

where

$M$  = amount of chlorine found,  $\mu\text{g}$ ; and

$V$  = volume of air sampled,  $\text{m}^3$ .

For different temperatures and atmospheric pressures, proper correction for air volume should be made.

**2.10 Effect on Storage** — The colour of sampled solutions is stable for at least 24 h if protected from direct sunlight, although the presence of certain interferences ( Fe III ) may cause slow colour change.



( Continued from page 2 )

**Methods of Sampling and Analysis Subcommittee, CDC 53 : 2**

*Convener*

SHRI P. K. YENNAWAR

*Representing*

National Environmental Engineering Research  
Institute ( CSIR ), Nagpur

*Members*

DR G. H. PANDYA ( *Alternate to*  
Shri P. K. Yennawar )

SHRI A. L. AGGARWAL

National Institute of Occupational Health  
( ICMR ), Ahmadabad

SHRI C. B. RAIYANI ( *Alternate* )

SMT J. M. DESHPANDE

Municipal Corporation of Greater Bombay

DR V. S. GUPTA

The Fertilizer ( Planning & Development ) India  
Ltd, Sindri

SHRI S. B. SINHA ( *Alternate* )

SHRI S. C. KALE

Directorate General Factory Advice Service and  
Labour Institutes, Bombay

SHRI S. K. DANGWAL ( *Alternate* )

DR D. N. KELKAR

Bhabha Atomic Research Centre, Bombay

SHRI S. G. KRISHNAN

National Organic Chemical Industries Ltd, Bombay

SHRI M. M. LAL

Industrial Toxicology Research Centre ( CSIR ),  
Lucknow

DR P. N. VISHWANATHAN ( *Alternate I* )

DR J. L. KAW ( *Alternate II* )

SHRI S. K. MAIRA

S. F. India Ltd, Calcutta

SHRI S. P. MENE ( *Alternate* )

SHRI S. D. MAKHIJANI

Central Board for the Prevention and Control of  
Water Pollution, New Delhi

DR H. S. MATHUR ( *Alternate* )

SHRI C. V. RAMASWAMY

Hindustan Petroleum Corporation Ltd, Bombay

SHRI S. N. CONTRACTOR ( *Alternate* )

DR N. C. SAHA

Indian Oil Corporation Ltd, New Delhi

SHRI C. V. RAMA MURTHY ( *Alternate* )

DR J. K. SINHA

Central Mining Research Station ( CSIR ),  
Dhanbad

PROF R. D. SRIVASTAVA

Indian Institute of Technology, Kanpur

DR M. S. VAIDYA

The Dharamsi Morarji Chemical Co Ltd, Bombay

SHRI H. B. SINGH ( *Alternate* )

SHRI S. K. VELINKER

National Mineral Development Corporation Ltd,  
Hyderabad

SHRI G. S. R. K. RAO ( *Alternate* )

# **BUREAU OF INDIAN STANDARDS**

## **Headquarters:**

Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI-110002

Telephones : 331 01 31  
331 13 75

Telegrams : Manaksanstha  
(Common to all Offices)

## **Regional Offices:**

|  | <i>Telephone</i>       |
|--|------------------------|
| Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110002        | 331 01 31<br>331 13 75 |
| * Eastern : 1/14 CIT Scheme VII M, V.I.P. Road, Maniktola, CALCUTTA 700054 | 37 86 62               |
| Northern : SCO 335-336, Sector 34-A, CHANDIGARH 160 022                    | 60 38 43               |
| Southern : C.I.T. Campus, IV Cross Road, MADRAS 600113                     | 235 23 15              |
| † Western : Manakalaya, E9 MIDC, Marol, Andheri (East), BOMBAY 400093      | 832 92 95              |

## **Branch Offices:**

|  |            |
|--|------------|
| 'Pushpak', Nurmohamed Shaikh Marg, Khanpur, AHMADABAD 380001                       | 30 13 48   |
| ‡ Peenya Industrial Area, 1st Stage, Bangalore-Tumkur Road,,<br>BANGALORE 560058   | 39 49 55   |
| Gangotri Complex, 5th Floor, Bhadbhada Road, T.T. Nagar, BHOPAL 462003             | 55 40 21   |
| Plot No. 21 Satyanagar, BHUBANESHWAR 751007  | 40 36 27   |
| Kalaikathir Building, 6/48 Avanashi Road, COIMBATORE 641037                        | 21 01 41   |
| Plot No. 43, Sector 16 A, Mathura Road, FARIDABAD 121001                           | 8-28 88 01 |
| Savitri Complex, 116 G.T. Road, GHAZIABAD 201001                                   | 8-71 19 96 |
| 53/5 Ward No. 29, R.G. Barua Road, 5th By-lane, GUWAHATI 781003                    | 54 11 37   |
| 5-8-56C L.N. Gupta Marg, Nampally Station Road, HYDERABAD 500001                   | 20 10 83   |
| R 14 Yudhister Marg, C Scheme, JAIPUR 302005                                       | 38 13 74   |
| 117/418 B Sarvodaya Nagar, KANPUR 208005   | 21 68 76   |
| Seth Bhawan, 2nd Floor, Behind Leela Cinema, Naval Kishore Road,<br>LUCKNOW 226001 | 23 89 23   |
| Patiputra Industrial Estate, PATNA 800013  | 26 23 05   |
| C/o Smt. Sunita Mirakhur,,<br>66 D/C Annexe, Gandhi Nagar, JAMMU TAWI 180004       | —          |
| T.C. No. 14/1421, University P.O., Palayam, THIRUVANANTHAPURAM 695034              | 6 21 17    |

## **Inspection Offices (With Sale Point):**

|   |          |
|---|----------|
| Pushpanjali, 1st floor, 205-A, West High Court Road, Shankar Nagar Square,<br>NAGPUR 440010 | 525171   |
| Institution of Engineers (India) Building 1332 Shivaji Nagar,<br>PUNE 411005                | 32 36 35 |

|   |           |
|---|-----------|
| * Sales Office is at 5 Chowringhee Approach, P.O. Princep Street,<br>CALCUTTA 700072      | 27 99 65  |
| † Sales Office is at Novelty Chambers, Grant Road, BOMBAY 400007                          | 309 65 28 |
| ‡ Sales Office is at 'F' Block, Unity Building, Narasimharaja Square,<br>BANGALORE 560002 | 22 39 71  |