Indian Standard

SPECIFICATION FOR APPARATUS FOR DETERMINING CONSTITUENTS OF FRESH CONCRETE

(Second Reprint FEBRUARY 1987)

UDU 666.972:620.168.3.05



© Copyright 1974

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

Indian Standard

SPECIFICATION FOR APPARATUS FOR DETERMINING CONSTITUENTS OF FRESH CONCRETE

Cement and Concrete Sectional Committee, BDC 2

Nadu

Hyderabad

Chairman

Representing

Dr H. C. VISVESVARAYA

Cement Research Institute of India, New Delhi

Members

DR A. S. BHADURI

National Test House, Calcutta

SHRIE, K. RAMACHANDRAN (Alternate)

SHRI A. K. CHATTERJI

Central Building Research Institute (CSIR). Roorkee

Public Works Department, Government of Tamil

DR S. S. REHSI (Alternate) DEPUTY CHIEF ENGINEER (BUILDINGS)

DEPUTY CHIEF ENGINEER (IRRIGATION & DESIGNS) (Alternate)

DR R. K. GHOSH (Alternate) DIRECTOR (CSMRS)

DEPUTY DIRECTOR (CSMRS)

(Alternate)

Central Road Research Institute (CSIR), New Delhi Central Water & Power Commission, New Delhi

Cement

SHRI K. H. GANGWAL

DIRECTOR

Hyderabad Alokudyog Services Ltd, New Delhi Indian Roads Congress, New Delhi

Asbestos

SHRI K. C. GHOSAL SHBI A. K. BISWAS (Alternate)

DR R. K. GHOSH

Brig Harish Chandra

SHRI G. R. MIRCHANDANI (Alternate) DR R. R. HATTIANGADI

SHRI P. J. JAGUS (Alternate) DR IQBAL ALI

IOINT DIRECTOR STANDARDS (B&S) DEPUTY DIRECTOR STANDARDS

(B&S) (Alternate)

SHRI S. B. JOSHI SHRI M. T. KANSE Engineer-in-Chief's Branch, Army Headquarters

Associated Cement Companies Ltd. Bombav

Engineering Research Laboratories, Hyderabad Research, Designs & Standards Organization, Lucknow

S. B. Joshi & Co Ltd, Bombay Directorate General of Supplies & Disposals

(Continued on page 2)

Products

Ltd.

@ Copyright 1974

INDIAN STANDARDS INSTITUTION

This publication is protected under the Indian Copyright Act (XIV of 1957) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

IS: 7325 - 1974

(Continued from page 1)

Members

SHRI S. L. KATHURIA SHRI S. R. KULKARNI

SHRI M. A. MEHTA SHRI O. MUTHACHEN

> SUPERINTENDING ENGINEER, 2ND CIRCLE (Alternate)

SHRI ERACH A. NADIRSHAH SHRI K. K. NAMBIAR

PROF G. S. RAMASWAMY

DR N. S. BHAL (Alternate) DR A. V. R. RAO

SHRI K. S. SRINIVASAN (Alternate)

SHRI G. S. M. RAO

SHRI T. N. S. RAO

SHRI S. R. PINHEIRO (Alternate

SECRETARY

DEPUTY SECRETARY (I) (Alternate)

SHRI R. P. SHARMA SHRI MOHINDER SINGH (Alternate)

SHRI G. B. SINGH SHRI C. L. KASLIWAL (Alternate)

SHRI J. S. SINGHOTA

SHRI T. C. GARG (Alternate)

Shri R. K. Sinha

SHRI K. A. SUBRAMANIAM SHRI P. S. RAMACHANDRAN (Alternate)

SHRI L. SWAROOP SHRI A. V. RAMANA (Alternate)

SHRI D. AJITHA SIMHA, Director (Civ Engg)

Roads Wing, Ministry of Shipping & Transport M. N. Dastur & Co (Private) Ltd, Calcutta Concrete Association of India, Bombay

Representing

Central Public Works Department

Institution of Engineers (India), Calcutta

In personal capacity ('Ramanalaya' 11 First Crescent Park Road, Gandhinagar, Adyar, Madras) Structural Engineering Research Centre (CSIR),

Roorkee

National Buildings Organization, New Delhi

Geological Survey of India, Nagpur

Gammon India Ltd, Bombay

Central Board of Irrigation & Power, New Delhi

Irrigation & Power Research Institute, Amritsar

Hindustan Housing Factory Ltd, New Delhi Beas Designs Organization, Nangal Township

Indian Bureau of Mines, Nagpur India Cements Ltd, Madras

Dalmia Cement (Bharat) Ltd, New Delhi

Director General, ISI (Ex-officio Member)

Secretary

SHRI Y. R. TANEJA

Deputy Director (Civ Engg), ISI

Instruments for Cement and Concrete Testing Subcommittee, BDC 2:10

Convener

DR IOBAL ALI

Engineering Research Laboratories, Hyderabad

Members

PROF B. M. AHUJA LALA G. C. DAS

SHRI T. P. EKAMBARAM DR R. K. GHOSH

SHRI K. L. SETHI (Alternate)

Indian Institute of Technology, New Delhi National Test House, Calcutta

Highways Research Station, Madras

Central Road Research Institute (CSIR), New Delhi

(Continued on page 10)

Indian Standard

SPECIFICATION FOR APPARATUS FOR DETERMINING CONSTITUENTS OF FRESH CONCRETE

O. FOREWORD

- **0.1** This Indian Standard was adopted by the Indian Standards Institution on 21 February 1974, after the draft finalized by the Cement and Concrete Sectional Committee had been approved by the Civil Engineering Division Council.
- 0.2 The Indian Standards Institution has already published a series of standards on methods of testing cement and concrete. It has been recognized that reliable and inter comparable test results can be obtained only with standard testing equipment capable of giving the desired level of accuracy. The Sectional Committee has, therefore, decided to bring out a series of specifications covering the requirements of equipment used for testing cement and concrete, to encourage its development and manufacture in the country.
- 0.3 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS: 2-1960*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard covers the requirements of apparatus for determining the proportions of cement, and fine and coarse aggregates of freshly mixed concrete where the nominal size of the largest aggregate does not exceed 40 mm.

2. WORKING PRINCIPLE

2.1 The test consists essentially in weighing the sample of concrete in air and in water, and then thoroughly washing it through two sieves to separate coarse and fine aggregates and to remove cement. The clean


^{*}Rules for rounding off numerical values (revised).

IS: 7325 - 1974

aggregates so recovered are then weighed in water and the mix proportions by weight are calculated from a knowledge of the specific gravities of the aggregates and cement.

3. DIMENSIONS

3.1 Dimensions of different components of the apparatus for determining constituents of fresh concrete shall be as detailed in Fig. 1. Where tolerances are not specifically mentioned in the relevant clauses, dimensions shall be considered nominal.




All dimensions in millimetres.

Fig. 1 5 kg Semi-Automatic Balance Reading to 0.5 g

4. PARTS AND ACCESSORIES

- 4.1 The apparatus shall consist of the parts given in 4.1.1 to 4.1.9.
- 4.1.1 Balance The balance shall be a semi-automatic one capable of weighing up to 5 kg to an accuracy of 0.5 g. The balance shall be provided with two counterpoises to obtain equilibrium, one for when an empty bucket is being weighed in air and the other for when it is being weighed whilst immersed in water (Fig. 1).

4.1.2 Bucket Shaped Containers — At least eight seamless bucket shaped cotainers of corrosion-resisting metal, such as, stainless steel, chromium or nickel plated steel sheet. 2.0 mm thick each 200 ± 5 mm diameter (internal) on top and 180 ± 5 mm deep having smooth sloping side and a rounded bottom to prevent the trapping of air when it is immersed, in accordance with Fig. 2. All these containers shall be of same weight in air correct to ± 0.25 g and each shall be clearly marked with the necessary correction figure to allow for the difference between its loss in weight when immersed in water and the weight of second counterpoise.




All dimensions in millimetres.

Fig. 2 Bucket Shaped Container

IS: 7325 - 1974

4.1.3 Tank — The tank shall be large enough to receive the bucket shaped container easily and made of corrosion resisting metal, such as stainless steel, chromium or nickel plated steel sheet 2.0 mm thick. tank shall be approximately 300 mm internal diameter and 300 mm deep in accordance with Fig. 3. This shall have an overflow spout in such a position that the rim of a bucket hung from the balance is completely immersed when the tank is full. The spout shall preferably be at a distance of 5 mm from the top of the tank. The tank shall be connected by a tap and flexible pipe to a subsidiary tank, so that the water level in the main tank can be raised or lowered without disturbance by-altering the level of the subsidiary tank. The tap shall be preferably at a distance of 20 mm from the top of the tank, so that in the lowest position of the subsidiary tank when the tap is open the level of water is below the lip of a bucket hanging on the balance. A haffle plate of corrosion resisting metal, such as stainless steel, chromium or nickel plated steel sheet (approximately 1.6 mm thick, 70 mm deep and 100 mm long) shall be provided inside the main tank opposite the inlet of the tap extending from the top of the tank to a position 50 mm below the centre of connection of the tap.



All dimensions in millimetres.

Fig. 3 Main Tank

- 4.1.4 Subsidiary Tank The subsidiary tank shall be of suitable size and made of corrosion resisting metal, such as stainless steel, chromium or nickel plated steel sheet. The bottom of this tank shall be connected to the tap of the main tank by a flexible pipe. The position of the subsidiary tank shall be movable so that it can be lowered or raised from the position of the tap on the main tank as necessity arises.
- 4.1.5 Sieve Two nesting sieves 450 mm diameter; one IS Sieve 4.75-mm (100 mm in depth) and other IS Sieve 150-microns (300 mm in depth) supported in the sieve frame by an adequately rigid spider of non-corrodible material, such as brass in accordance with Fig. 4. The sieves shall conform to IS: 460-1962*.
- 4.1.6 Funnel A funnel approximately 500 mm in diameter at the top, 150 mm diameter at the bottom and 250 mm deep, made of 1.6 mm thick corrosion resistant metal sheet, such as, stainless steel, chromium or nickel plated steel on a sample bucket in accordance with Fig. 5 shall be provided to facilitate the washing of materials from the large sieves into the buckets without loss of any material.
- 4.1.7 Hose A hose fitted with a nozzle giving a fairly fine strong spray of water shall be provided for washing of the cement from the aggregates over the surface of the sieves.
- 4.1.8 Stirring Rod A mild steel stirring rod of 15 mm diameter and 300 mm long shall be provided.
- 4.1.9 Weight Box A standard weight box of capacity 5 kg shall be provided.


5. MATERIALS

5.1 The materials that shall be used for making the different items of the apparatus are given in Table 1.

6. MARKING


- 6.1 The following intermation shall be clearly and indelibly marked on each component of the apparatus in a way that it does not interfere with the performance of the apparatus:
 - a) Name of manufacturer or his registered trade-mark or both, and
 - b) Date of manufacture.

^{*}Specification for test sieves (revised)



All dimensions in millimetres.

Fig. 4 Nested Sieves



All dimensions in millimetres.

Fig. 5 Funnel for Transfer of Material

6.1.1 The apparatus may also be marked with the ISI Certification Mark.

NOTE—The use of the ISI Certification Mork is governed by the provisions of the Indian Standards Institution (Certification Marks) Act and the Rules and Regulations made thereunder. The ISI Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well-defined system of inspection, testing and quality control which is devised and supervised by ISI and operated by the producer. ISI marked products are also continuously checked by ISI for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

TABLE 1 MATERIALS FOR DIFFERENT ITEMS OF APPARATUS FOR DETERMINING CONSTITUENTS OF FRESH CONCRETE

(Clause 5.1)

SL No.	ITEM	Material	SPECIAL REQUIRE- MENT, IF ANY	RECOMMENDED Indian Standard, IF ANY
(1)	(2)	(3)	(4)	(5)
i)	Bucket includ- ing handle	Non-corrodible metal not attacked by cement, such as stain- less steel, chromium or nickel plated steel	Smooth finish (inside and outside)	7
ii)	Tank	Non-corrodible metal not attacked by cement, such as stain- less steel, chromium or nickel plated steel	Smooth inter- nal finish	_
iii)	Spider	Brass	_	IS: 292-1961*
iv)	Stand for sample bucket	Cast iron	-	IS: 210-1970†

^{*}Specification for brass ingots and castings (revised).
†Specification for grey iron castings (second revision).

(Continued from page 2)

Members

Representing

SHRI H. K. GUHA

All India Instruments Manufacturers & Dealers Association, Bombay

SHRI V. K. VASUDEVAN (Alterante)

SHRI P. J. JAGUS Associated Cement Companies Ltd, Bombay

SHRI D. A. WADIA (Alternate)

SHRI M. R. JOSHI

Ministry of Defence, R & D

SHRI K. A. SIRASKAR (Alternate)

Cement Research Institute of India, New Delhi SHRI V. KANNAN

SHRI M. V. RANGA RAO (Alternate)

Indian Institute of Technology, Bombay PROF C. K. RAMESH

DR R. S. AYYAR (Alternate) SHRI P. V. SUBBA RAO

Andhra Scientific Co Ltd, Masulipatam

SHRI Y. S. NARAYANA (Alternate)

DR S. S. REHSI Central Building Research Institute (CSIR), Roorkee

SHRI J. P. KAUSHISH (Alternate) SHRI H. C. VERMA Associated Instrument Manufacturers (India) Pvt Ltd, New Delhi

SHRI A. V. SHASTRI (Alternate)



157 INDIAN STANDARDS INSTITUTION

Headquarters:					
Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110002					
Telephones: 3 31 01 31, 3 31 13 75 Telegrams: Ma	naksanstha				
Regional Offices :	Telephone				
*Western ; Manakalaya, E9 MIDC, Marol, Andheri (East). BOMBAY 400093	6 32 92 95				
†Eastern: 1/14 C. I. T. Scheme VII M, V. I. P. Road, Maniktola, CALCUTTA 700054	36 24 99				
Northern: SCO 445-446, Sector 35-C CHANDIGARH 160036	{2 18 43 3 16 41				
Southern: C. I. T. Campus, MADRAS 600113	\$\\\\ 41 24 42 \\\\ 41 25 19 \\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\				
	141 29 16				
Branch Offices:					
Pushpak, Nurmohamed Shaikh Marg, Khanpur, AHMADABAD 380001	{2 63 48 2 63 49				
'F' Block, Unity Bldg, Narasimharaja Square, BANGALORE 560002	22 48 05				
Gangotri Complex, 5th Floor, Bhadbhada Road, T. T. Nagar BHOPAL 462003	, 6 27 16				
Plot No. 82/83, Lewis Road, BHUBANESHWAR 751002	5 36 27				
53/5 Ward No. 29, R. G. Barua Road, 5th Byelane, GUWAHATI 781003	-				
5-8-56C L. N. Gupta Marg. (Nampally Station Road), HYDERABAD 500001	22 10 83				
R14 Yudhister Marg, C Scheme, JAIPUR 302005	{6 34 71 6 98 32				
117/418B Sarvodaya Nagar, KANPUR 208005	{21 68 76 21 82 92				
Patliputra Industrial Estate, PATNA 800013	6 23 05				
Hantex Bldg (2nd Floor), Rly Station Road, TRIVANDRUM 695001	52 27				
Inspection Office (With Sale Point):					
Institution of Engineers (India) Building, 1332 Shivaji Nac PUNE 410005	par, 5 24 35				
*Sales Office in Bombey is at Novelty Chambers, Grant Road, Bombey 400007	89 65 28				
†Sales Office in Calcutte is at 5 Chowringhee Approach, P. O. Prince Street, Calcutte 700072	p 27 88 00				