UCM, TG119 0097

Clinic: RadOnc UChicago Medicine

5758 S Maryland Avenue

Chicago IL 60637

Plan: 4 CShape50-D4

Planned: 11/20/2019 2:30 PM Ji Li Measured: 11/21/2019 6:46 PM planning


Comments:

Treatment Summary

Radiation Device: TrueBeamSN1203

Temperature: 22.0 °C Reference: Planned Dose

Beam	Gantry	Energy	Daily corr	Norm	Det within acc	eptance	Median
			factor	dose	Dev DTA	γ-index	dose dev
Composite				261 cGy	100.0%99.3%	99.5%	0.1%
Field 1	210° to 150°	6 MV	1.000	105 cGy	99.0% 99.6%	100.0%	0.2%
Field 2	150° to 210°	6 MV	1.000	111 cGy	99.8% 99.2%	99.3%	-0.0%
Field 3	210° to 150°	6 MV	1.000	115 cGy	100.0%99.3%	99.6%	0.2%



Printed: 12/2/2019 10:22 AM Page 1(4)


UCM, TG119 0097

Gai	mma	a Index	k Evalı	uations	s —						
DT	ГΑ					Pass	s Rates	[%]			Reference: Planned dose
	mm	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
4.5	mm	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
4.0	mm	99.3	99.8 99.8	99.8 99.8	100.0 100.0	100.0 100.0	100.0	100.0	100.0 100.0	100.0	100.0 100.0
3.5 3.0 2.5 2.0	mm	99.2 98.4	99.6 99.5	99.6 99.5	99.7	100.0	100.0 100.0	100.0 100.0	100.0	100.0 100.0	100.0
E 25	mm	96.4 96.9	99.0	99.5	99.7 99.7	99.7	99.8	100.0	100.0	100.0	100.0
5 2.3	mm	94.2	97.4	98.7	99.5	99.7	99.8	99.8	100.0	100.0	100.0
1.5	mm	90.4	94.7	97.0	99.2	99.5	99.5	99.7	100.0	100.0	100.0
	mm	83.7	91.8	94.9	97.5	98.8	99.5	99.7	99.7	100.0	100.0
	mm	66.9	82.2	92.4	95.9	97.7	98.7	99.7	99.7	100.0	100.0
5.0	mm	99.7	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
4.5	mm	99.7	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	99.5	99.7	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	99.0	99.7	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0
3.0	mm	98.5	99.0	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0
3.0 2.5 2.0	mm	97.0	98.5	99.5	100.0	100.0	100.0	100.0	100.0	100.0	100.0
■ 2.0	mm	95.5	97.7	98.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	91.9	95.3	97.1	99.2	100.0	100.0	100.0	100.0	100.0	100.0
	mm	82.7	88.4	93.6	96.5	98.3	98.8	99.5	100.0	100.0	100.0
0.5	mm	59.7	76.7	85.1	91.3	95.6	97.5	98.2	98.8	99.0	99.3
5.0	mm	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	99.8	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	99.5	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	99.5	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
		99.1	99.6	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
2.5	mm	97.5	99.5	99.6	99.8	100.0	100.0	100.0	100.0	100.0	100.0
3.0 2.5 2.0	mm	94.8	98.0	99.1	99.3	99.6	99.8	100.0	100.0	100.0	100.0
1.5	mm	90.3	94.4	97.8	98.7	99.1	99.6	99.6	99.8	99.8	100.0
	mm	78.6	89.2	94.8	97.7	98.4	99.1	99.5	99.8	99.8	99.8
0.5	mm	56.7	75.2	85.6	93.2	96.1	98.0	98.7	99.6	99.8	99.8
5.0	mm	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0	100.0
	mm	99.8	99.8	99.8	100.0	100.0	100.0	100.0	100.0	100.0	100.0
3.5	mm	99.5	99.6	99.8	99.8	100.0	100.0	100.0	100.0	100.0	100.0
		98.8	99.1	99.5	99.8	100.0	100.0	100.0	100.0	100.0	100.0
3 2.5	mm	97.9	98.4	98.9	99.8	100.0	100.0	100.0	100.0	100.0	100.0
3.0 2.5 2.0	mm	96.6	97.9	98.8	99.6	100.0	100.0	100.0	100.0	100.0	100.0
1.5	mm	94.1	96.8	97.9	99.1	99.6	100.0	100.0	100.0	100.0	100.0
1.0	mm	85.0	92.7	95.5	97.9	99.3	99.8	100.0	100.0	100.0	100.0
0.5	mm	64.9	77.9	89.3	94.7	97.0	98.8	99.3	100.0	100.0	100.0
		0.5 %	1.0 %	1.5 %	2.0 %	2.5 %	3.0 %	3.5 %	4.0 %	4.5 %	5.0 %
							e Devia				
						200					

Parameter	Selected Detectors	Δ Dose	Δ Dist	Acceptance Limits
Dose Deviation	Dose from 10% to 500%	n.a.	n.a.	90% within ±5.0%
Dist to Agreement	Gradient >= 1%/mm	n.a.	n.a.	90% with DTA <= 3.0 mm
Gamma Index	Dose from 10% to 500%	±2.0%	2.0 mm	95% with gamma <= 1

Printed: 12/2/2019 10:22 AM Page 2(4)

UCM, TG119 0097



UCM, TG119 0097

