

GeneLab SOP for Quant-iT PicoGreen dsDNA quantification of Illumina sequencing libraries	Document No.:	GL-SOP-6.2
	Version:	1.0
	Created:	09 18 2018
	Last revised:	02 24 2020
	Last revised by:	Yi-Chun Chen

Purpose/Scope:

This procedure lists the steps for dsDNA quantification of sequencing libraries using the Quant-iT™ PicoGreen™ dsDNA Assay Kit.

Equipment:

1. 96-well flat bottom optical plate
2. Vortex mixer that can hold 96 well plates
3. Centrifuge that can hold 96 well plates

Reagents:

1. Quant-iT™ PicoGreen™ dsDNA Assay Kit – P11496 (Kit contains Quant-iT PicoGreen dsDNA reagent, 20X TE, and DNA standard).

For 100 reactions to quantify up to 26 samples use:

55 uL dye

30 mL 1X TE

300 ul each of diluted samples

300 ul each of the standards


Procedure:

1. Thaw all components at room temperature for 15 min. Protect dye from light.
2. Prepare 1X TE by mixing 1.5 mL 20X TE to 28.5 mL water.
3. Prepare 1X dye (200-fold dilution) by mixing the following:
 - a. For 100 reactions: 55 uL dye + 10.945 mL 1X TE (Final volume: 11mL)
4. Prepare DNA standard (2 uG/mL) from stock (100 ug/mL). Adjust dilution if using stock with different concentration.
 - a. Standard 1: Mix 10 uL stock DNA + 490 uL of 1X TE
5. Make 6 DNA standard as follows (2X serial dilution). Vortex between each standard.
 - b. Standard 2: 400 uL of 2 ug/mL DNA stock + 400 uL of TE (1 ug/mL)
 - c. Standard 3: 500 uL of STD 2 + 500 uL of TE (500 ng/mL)
 - d. Standard 4: 500 uL of STD 3 + 500 uL of TE (250 ng/mL)
 - e. Standard 5: 500 uL of STD 4 + 500 uL of TE (125 ng/mL)
 - f. Standard 6: 500 uL of STD 5 + 500 uL of TE (62.5 ng/mL)
 - g. Standard 7: 500 uL of STD 6 + 500 uL of TE (31.25 ng/mL)
 - h. Standard 8: 500 uL of TE (Blank)
6. Dilute all test samples 100-fold in 1X TE, mix well.
 - a. 4 uL of samples + 396 uL of 1X TE
7. Load 3X 100 uL of standards 3-8 and samples onto the optical plate. See below for plate layout.


8. Add 100uL of the working dye to each well. Mix gently with pipette.
9. Seal plate with aluminum foil, vortex until well mixed.
10. Centrifuge the plate to collect droplets. Measure after 5 min incubation.

Measure with plate reader (excitation ~480nm, emission ~520nm)


1. Turn on SpectraMax i3x and connect to the computer.
2. Open SoftMax Pro 6 on the computer.
3. Click on the plate icon -> double click on "GL_PicoGreen_TEMPLATE.sda" to open template-> Right click on Expt1 and give the experiment a new name -> Use the "save as" function to save to a new file.



4. Check to make sure all the parameters are correct.



- Click on "Template Editor" to view or change plate layout.



- Click "Read" to read the plate.
- If reading more than 1 plate, click on "New Experiment" at the top. Change the experiment name to distinguish between different plates.

Sample plate layout

STD 3 (500ng/mL)	STD 3 (500ng/mL)	STD 3 (500ng/mL)	STD 4 (250ng/mL)	STD 4 (250ng/mL)	STD 4 (250ng/mL)	STD 5 (125ng/mL)	STD 5 (125ng/mL)	STD 5 (125ng/mL)	STD 6 (62.5ng/mL)	STD 6 (62.5ng/mL)	STD 6 (62.5ng/mL)
STD 7 (62.5ng/mL)	STD 7 (62.5ng/mL)	STD 7 (62.5ng/mL)	STD 8 Blank (TE)	STD 8 Blank (TE)	STD 8 Blank (TE)	Sample 1	Sample 1	Sample 1	Sample 2	Sample 2	Sample 2
Sample 3	Sample 3	Sample 3	Sample 4	Sample 4	Sample 4	Sample 5	Sample 5	Sample 5	Sample 6	Sample 6	Sample 6
Sample 7	Sample 7	Sample 7	Sample 8	Sample 8	Sample 8	Sample 9	Sample 9	Sample 9	Sample 10	Sample 10	Sample 10
Sample 11	Sample 11	Sample 11	Sample 12	Sample 12	Sample 12	Sample 13	Sample 13	Sample 13	Sample 14	Sample 14	Sample 14
Sample 15	Sample 15	Sample 15	Sample 16	Sample 16	Sample 16	Sample 17	Sample 17	Sample 17	Sample 18	Sample 18	Sample 18
Sample 19	Sample 19	Sample 19	Sample 20	Sample 20	Sample 20	Sample 21	Sample 21	Sample 21	Sample 22	Sample 22	Sample 22
Sample 23	Sample 23	Sample 23	Sample 24	Sample 24	Sample 24	Sample 25	Sample 25	Sample 25	Sample 26	Sample 26	Sample 26