

Standard Name21 of 31

↑ ADAT AluminaS20 ↓

Visual Hue Request: YELLOW/REDDISH

New Std	Retrieve Std	Store Std	List Std
---------	--------------	-----------	----------

Instrument Setup:%R LAV SCI UV Cal

Date: 19-Mar-24Time: 19:41:47

Batch Name3 of 3


↑ 9-FW24-FI-00867A .137615 DOM2400579 ↓

New Bat	Retrieve Bat	Store Bat	List Bat
---------	--------------	-----------	----------

Instrument Setup: %R LAV SCI UV Cal

Date: 16-May-24Time: 10:13:24

DeltaD65 / 10F11 / 10



Single/All	Zoom In	Zoom Out	CMC Plot	Print Form
------------	---------	----------	----------	------------

DataColor TOOLS 2.8.1S/N NO. 8823242

Standard %R

400-450nm	60.18	57.66	53.52	49.20	49.02	45.46
460-510nm	48.70	57.74	61.57	61.41	60.56	58.43
520-570nm	56.14	56.19	56.57	55.77	55.16	56.73
580-630nm	59.20	60.44	61.32	61.91	61.31	60.10
640-690nm	59.84	61.97	67.60	75.25	82.99	87.43
700nm	91.00					

CMC DE Pass/Fail Tolerance: 0.80

I:c Ratio: (2.00 1.00)

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
60012609 240319-FW24-FI-00867A .137615 DOM2400579	D65 10 Deg	-1.13	0.33	0.51	-0.42	0.50	-0.45	0.79		Warn	too dark	too bright	too red
60012609 240319-FW24-FI-00867A .137615 DOM2400579	F11 10 Deg	-1.03	-0.49	0.34	-0.38	0.25	0.79	0.91	0.83	Fail	too dark	too bright	too green