

Standard Name

367 of 415

DataColor TOOLS 2.8.1 S/N NO. 8823242

AE6M Focus OliveF21

Visual Hue Request: GREEN/YELLOWISH

New Std Retrieve Std Store Std List Std

Instrument Setup: %R LAV SCI UV Cal

Date: 14-Feb-22 Time: 11:33:23


Batch Name 7 of 7

62733223 POA12412A/E0027 182582. DO

New Bat Retrieve Bat Store Bat List Bat

Instrument Setup: %R LAV SCI UV Cal

Date: 18-Dec-24 Time: 15:26:25



Standard %R

400-450nm	11.16	9.05	7.72	6.59	6.02	5.93
460-510nm	6.03	6.51	7.75	9.04	9.86	10.33
520-570nm	11.02	11.70	11.73	11.50	11.01	10.62
580-630nm	10.43	10.36	10.39	10.44	10.78	11.61
640-690nm	13.37	16.10	19.70	24.02	29.09	33.60
700nm	39.67					

CMC DE Pass/Fail Tolerance: 0.80

I:c Ratio: (2.00 1.00)

Batch Name

62733223 POA12412A/E0027 182582. DOM2401813
62733223 POA12412A/E0027 182582. DOM2401813

Single/All Zoom In Zoom Out CMC Plot Print Form

III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments
D65 10 Deg	-0.47	0.48	0.64	-0.25	0.42	-0.62	0.79		Warn	too dark	too bright
F11 10 Deg	-0.32	0.59	0.78	-0.17	0.43	-0.77	0.90	0.23	Fail	too dark	too bright