

CALIBRATION REPORT

CUSTOMER : NACM - INDAH WATER KONSORTIUM SDN BHD
: NO. 12-2A, JALAN 15/1F, SEKSYEN 15, 43650 B.B.BANGI,
SELANGOR

CERTIFICATE : TMSB/SV/0158
NUMBER
DATE OF ISSUE : 12/2/2026
PAGE : 1 OF 1

DEVICE UNDER TEST (UUT)

Instrument Name : Safety Valve
Manufacturer : -
Serial Number/PMT : TMSB/00151
Set Pressure : 13 BAR
Model No : N/A
Tag No : N/A
Received Condition/Returned Condition : Good Physical Condition
Calibration Location : NACM - INDAH WATER KONSORTIUM SDN BHD
Calibration Date : 12/2/2026
Due Date : 12/2/2027

The user should be aware that any number of factors may cause this instrument to drift out of calibration before the specified calibration interval has expired.

Instrument when received : ITEM RECEIVED IN GOOD CONDITION
Instrument when return : The item has been calibrated. The result as below
Calibration Procedure : TMCP

Calibration Standard(s) Used

Calibration was performed using the comparison method between a traceable master instrument and the unit under test (UUT). Reading were taken at specified test points and the deviation was recorded

MASTER INSTRUMENT INFORMATION

Instrument Type:	Certificate :	No. Serial number :	Due Date:	Traceability:
1 Pressure Gauge	VNA/A3880/25	PR2407101	18 Mar 2026	NM

SET PRESSURE (bar)	ACTUAL PRESSURE (bar)	CORRECTION (bar)	TOLERANCE (% of set pressure)
13	13	0	0


RESULT : PASS

- 1) The expanded uncertainty of measurement associated with the calibration is ± 0.3 bar estimated at a confidence level of approximately 95 % with a coverage factor of $k=2$.
- 2) The user should determine the suitability of the instrument for its intended use

Calibrated by:


Mohd Syukri Yusof
Calibration Officer

Approved by:


Shalihin Idris
Approving Officer

