

ACCEPTANCE CRITERIA FOR INSPECTION:

- 1. INSPECT FOR LOOSE DEBRIS OR PARTICLES.
- 2. INSPECT FOR UNIFORM SURFACE CONDITION.
- 3. INSPECT FOR COLOR, DISCOLORATION, HOLES OR TEARS.

3.0 FABRICATE USING 3D SOLID MODEL FILE PER REVISION BLOCK. ALL DIMENSIONS AND FEATURES NOT SPECIFIED ON PRINT ARE TO BE DERIVED FROM 3D MODEL FILE. INSPECT PART PER PART PRINT.

2.0 MATERIAL COLOR: CLEAR

1.0 MATERIAL TYPE: SILICONE

NOTES: UNLESS OTHERWISE NOTED

REVISIONS			
REV	ECO NUMBER	RELEASE DATE	DESCRIPTION OF CHANGE
0	040	06/19/2023	INITIAL RELEASE
1	063		UPDATED TO SHOW FLAT

UNLESS OTHERWISE SPECIFIED:

DIMENSIONS ARE IN INCHES

TOLERANCES:

ONE PLACE DECIMAL ±.100

TWO PLACE DECIMAL ±.010


THREE PLACE DECIMAL ±.005

THIRD ANGLE PERSPECTIVE:

PROPRIETARY AND CONFIDENTIAL

THE INFORMATION CONTAINED IN THIS DRAWING IS THE SOLE PROPERTY OF STARLING MEDICAL. ANY REPRODUCTION IN PART OR AS A WHOLE WITHOUT THE WRITTEN PERMISSION OF NEOTECH PRODUCTS IS PROHIBITED.

	NAME	DATE
DRAWN BY	D.SALDANA	02/07/2023
APPROVED BY	D. HENDRICKS	06/15/2023
SEE ECO RECORDS FOR ALL APPROVALS		
APPLICABLE STANDARDS: INTERPRET DIMENSIONS AND TOLERANCES PER ASME Y14.5-2009.		
COMMENTS:		



TITLE:

MOLDED, URINDX, O-RING, PERIMETER SEAL

SIZE	DWG. NO.	REV
B	DW-01113	1

SCALE:

SHEET 1 OF 1