

UUT Report

Station ID: IT-W-7303101
Serial Number: NONE
Date: venerdì 23 giugno 2023
Time: 14:10:16
Operator: administrator
Execution Time: 40.1987146 seconds
Number of Results: 30

Error, Error executing substep 'Run'.
Error 56 occurred at IPSES_Modbus_x86.lvlibp:RTU Data Unit.lvclass:Read ADU Packet.vi:5030001

UUT Result: Possible reason(s):

LabVIEW: (Hex 0x38) The network operation exceeded the user-specified or system time limit.
LabVIEW: (Hex 0x38) The network operation exceeded the user-specified or system time limit.
[Error Code: 56, User-defined error code.]

Serial Number: 4294967295
Type Designator: B3-D4MB-111A
FW Version: 0.16.1
Power Supply: Zera

Begin Sequence: MainSequence
(C:\Users\itlavit1\OneDrive - ABB\LabRnD_Shared\TestStand\Sequences\EQ Meter\New Sequences\Energy Registers Power Fail.seq)

DUT info	
Status:	Done
Serial Number:	4294967295
Type Designator:	B3-D4MB-111A
FW Version:	0.16.1

Power_ON_generator	
Status:	Passed
Module Time:	19.8917344

Begin Sequence: Power_ON_generator
(C:\Users\itlavit1\OneDrive - ABB\LabRnD_Shared\TestStand\Sequences\Utilities\CommonUtilitySequencesEqRed.seq)

Zera ON	
Status:	Passed
Module Time:	14.7026843

End Sequence: Power_ON_generator

End Sequence: MainSequence

End UUT Report