

UUT Report

Station ID: IT-W-7303101  
Serial Number: NONE  
Date: martedì 11 aprile 2023  
Time: 14:11:09  
Operator: ITLAVIT1  
Execution Time: 00:02:55.011  
Number of Results: 32  
UUT Result: Failed  
Failure Chain:

Step	Sequence	Sequence File
<a href="#">PMD-1145</a>	MainSequence	PMD-1145.seq
<a href="#">Numeric Limit Test</a>	PMD-1145	PMD-1145.seq

Begin Sequence: MainSequence  
(C:\Users\itlavit1\OneDrive - ABB\LabRnD\_Shared\TestStand\Sequences\M4M\PMD-1145.seq)

DUT recap	
Status:	Done
TestResults/Data:	
Serial Number:	"
Type Designator:	"
FW Version:	"
PMD-1145	
Status:	Failed
Module Time:	172.466

Begin Sequence: PMD-1145  
(C:\Users\itlavit1\OneDrive - ABB\LabRnD\_Shared\TestStand\Sequences\M4M\PMD-1145.seq)

Numeric Limit Test	
Status:	Passed
Numeric:	0.0001
Limits:	
Low:	-2
High:	2
Comparison Type:	GELE(>= <=)
Numeric Limit Test	
Status:	Failed
Numeric:	8.6317
Limits:	
Low:	9
High:	11

Comparison Type:	GELE(>= <=)
Numeric Limit Test	
Status:	Failed
Numeric:	17.5127
Limits:	
Low:	18
High:	22
Comparison Type:	GELE(>= <=)
Configure Analog out 2	
Status:	Failed
Numeric Limit Test	
Status:	Passed
Numeric:	4.2082
Limits:	
Low:	2
High:	6
Comparison Type:	GELE(>= <=)
Numeric Limit Test	
Status:	Failed
Numeric:	16.8018
Limits:	
Low:	12
High:	15
Comparison Type:	GELE(>= <=)
Numeric Limit Test	
Status:	Passed
Numeric:	20.3223
Limits:	
Low:	18
High:	22
Comparison Type:	GELE(>= <=)

**End Sequence: PMD-1145**

**End Sequence: MainSequence**

**End UUT Report**

---