Test Software P/N 4102483

Tested Part Number LA1054096

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,7 (Setpoint 75) [16m 16	ADC Cnts	LE 47011	46037	1 1
37	0020.0580 Pos Flow Cal: dP2 at 87,7 (Setpoint 85) [16m 16	ADC Cnts	LE 47942	47012	→
38	0020.0590 Pos Flow Cal: dP2 at 97,7 (Setpoint 95) [16m 16	ADC Cnts	LE 48819	47943	√
39	0020.0600 Pos Flow Cal: dP2 at 107,3 (Setpoint 105) [16m	ADC Cnts	LE 49614	48820	1
40	0020.0610 Pos Flow Cal: dP2 at 117,3 (Setpoint 115) [16m	ADC Cnts	LE 50337	49616	→
41	0020.0620 Pos Flow Cal: dP2 at 127,5 (Setpoint 125) [16m	ADC Cnts	LE 51018	50338	→
42	0020.0630 Pos Flow Cal: dP2 at 137,5 (Setpoint 135) [16m	ADC Cnts	LE 51662	51019	→
43	0020.0640 Pos Flow Cal: dP2 at 147,3 (Setpoint 145) [16m	ADC Cnts	LE 52868	51664	→
44	0020.0650 Pos Flow Cal: dP2 at 167,0 (Setpoint 170) [16m	ADC Cnts	LE 54139	52869	→
45	0020.0660 Pos Flow Cal: dP2 at 188,0 (Setpoint 190) [16m	ADC Cnts	LE 65435	54140	→
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [lnHg/ADC	EQ 0,000517	0,000517	√
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (inHg	1,91 to 4,09	2,59	√
48	0020.0920 Write Sensors Cal Table [20m 31s]	N/A	Pass	TRUE	√
49	0020.0970 Write Pprox Cal Table [20m 55s]	N/A	Pass	TRUE	√
50	0020.0935 02 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	√
51	0020.1030 Write Image Table [21m 15s]	N/A	Pass	TRUE	√
					$-\!\!\!\!+\!\!\!\!-\!\!\!\!\!-$

Test Started On 08/29/24 10:35:56

Elapsed Test Time 21m 15s

Status PASS

Serial Number TV119091018

Page

2 **of** 2