

Test Software P/N 4102483Tested Part Number 1040000Test Software Name Trilogy FSA (REPAIR TEST)Software Revision 24.0.0.0Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0021.0007 OBM O2 Sensor Heater On (NO_O2_ERROR) [0s]	N/A	Pass	TRUE	✓	
2	0021.0010 O2 Press Sensor Cal P1@ 25,25 & 75,10 PSI: Slope	ADC Cnts	0,03010 to 0,0510	0,03327	✓	
3	0021.0020 O2 Press Sensor Cal P1@ 25,25 & 75,10 PSI: Inter	PSI	-25,00 to -10,00	-13,36	✓	
4	0021.0030 O2 Press Sensor Cal P1@ 25,25 & 75,10 PSI: Zero	ADC Cnts	320 to 490	408	✓	
5	0021.0120 Neg Flow Cal: dP2 at -48,2 (Setpoint 45) [5m 52s]	ADC Cnts	GE 101	119	✓	
6	0021.0130 Neg Flow Cal: dP2 at -32,8 (Setpoint 35) [5m 52s]	ADC Cnts	GE 120	361	✓	
7	0021.0140 Neg Flow Cal: dP2 at -23,0 (Setpoint 25) [5m 52s]	ADC Cnts	GE 362	531	✓	
8	0021.0150 Neg Flow Cal: dP2 at -15,0 (Setpoint 15) [5m 52s]	ADC Cnts	GE 532	731	✓	
9	0021.0160 Neg Flow Cal: dP2 at -5,0 (Setpoint 5) [5m 52s]	ADC Cnts	GE 732	1032	✓	
10	0021.0170 Neg Flow Cal: dP2 at -0,0 (Setpoint 0) [5m 52s]	ADC Cnts	GE 1034	1174	✓	
11	0021.0250 Raw Zero Flow: dP2 at 0,0 (Setpoint 0) [5m 52s]	ADC Cnts	GE 0	1174	✓	
12	0021.0300 Pos Flow Cal: dP2 at 5,0 (Setpoint 5) [9m 57s]	ADC Cnts	LE 1620	1315	✓	
13	0021.0310 Pos Flow Cal: dP2 at 15,0 (Setpoint 15) [9m 57s]	ADC Cnts	LE 1862	1621	✓	
14	0021.0320 Pos Flow Cal: dP2 at 26,0 (Setpoint 25) [9m 57s]	ADC Cnts	LE 2059	1863	✓	
15	0021.0330 Pos Flow Cal: dP2 at 38,0 (Setpoint 35) [9m 57s]	ADC Cnts	LE 2190	2060	✓	
16	0021.0340 Pos Flow Cal: dP2 at 47,0 (Setpoint 45) [9m 57s]	ADC Cnts	LE 2330	2191	✓	
17	0021.0350 Pos Flow Cal: dP2 at 57,0 (Setpoint 55) [9m 57s]	ADC Cnts	LE 2458	2331	✓	
18	0021.0360 Pos Flow Cal: dP2 at 68,0 (Setpoint 65) [9m 57s]	ADC Cnts	LE 2569	2459	✓	
19	0021.0370 Pos Flow Cal: dP2 at 77,3 (Setpoint 75) [9m 57s]	ADC Cnts	LE 2676	2570	✓	
20	0021.0380 Pos Flow Cal: dP2 at 87,5 (Setpoint 85) [9m 57s]	ADC Cnts	LE 2769	2677	✓	
21	0021.0390 Pos Flow Cal: dP2 at 97,0 (Setpoint 95) [9m 57s]	ADC Cnts	LE 2863	2770	✓	
22	0021.0400 Pos Flow Cal: dP2 at 107,0 (Setpoint 105) [9m 57s]	ADC Cnts	LE 2941	2864	✓	
23	0021.0410 Pos Flow Cal: dP2 at 117,8 (Setpoint 115) [9m 57s]	ADC Cnts	LE 3014	2942	✓	
24	0021.0420 Pos Flow Cal: dP2 at 127,7 (Setpoint 125) [9m 57s]	ADC Cnts	LE 3093	3015	✓	
25	0021.0430 Pos Flow Cal: dP2 at 137,5 (Setpoint 135) [9m 57s]	ADC Cnts	LE 3156	3094	✓	
26	0021.0440 Pos Flow Cal: dP2 at 147,5 (Setpoint 145) [9m 57s]	ADC Cnts	LE 3320	3158	✓	
27	0021.0450 Pos Flow Cal: dP2 at 172,5 (Setpoint 175) [9m 57s]	ADC Cnts	LE 3416	3321	✓	
28	0021.0460 Pos Flow Cal: dP2 at 187,8 (Setpoint 190) [9m 57s]	ADC Cnts	LE 3999	3417	✓	
29	0021.0500 O2 Positive Flow Cal: dP1 at 0,0 (SP 0 @80,7PSI)	ADC Cnts	LE 418	241	✓	
30	0021.0510 O2 Positive Flow Cal: dP1 at 4,3 (SP 5 @80,6PSI)	ADC Cnts	LE 817	418	✓	
31	0021.0520 O2 Positive Flow Cal: dP1 at 15,0 (SP 15 @79,3PSI)	ADC Cnts	LE 1057	818	✓	
32	0021.0530 O2 Positive Flow Cal: dP1 at 24,7 (SP 25 @78,2PSI)	ADC Cnts	LE 1267	1058	✓	
33	0021.0540 O2 Positive Flow Cal: dP1 at 34,0 (SP 35 @77,2PSI)	ADC Cnts	LE 1453	1268	✓	
34	0021.0550 O2 Positive Flow Cal: dP1 at 44,0 (SP 45 @76,0PSI)	ADC Cnts	LE 1629	1454	✓	
35	0021.0560 O2 Positive Flow Cal: dP1 at 54,3 (SP 55 @74,5PSI)	ADC Cnts	LE 1781	1630	✓	

Test Started On 07/10/24 12:08:55Serial Number TV017031585Elapsed Test Time 16m 32sStatus **PASS**

Page 1 of 2

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20