

Test Software P/N 4102483Tested Part Number 1040000Test Software Name Trilogy FSA (PM TEST)Software Revision 24.0.0.0Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0570 Neg Flow Verify: dP2 at -116,7 (Setpoint 120) [36m]	SLPM	-124,8 to -108,5	-121,3	✓	
37	0030.0580 Neg Flow Verify: dP2 at -107,2 (Setpoint 110) [36m]	SLPM	-115,0 to -99,4	-110,8	✓	
38	0030.0590 Neg Flow Verify: dP2 at -97,5 (Setpoint 100) [36m]	SLPM	-104,9 to -90,1	-99,8	✓	
39	0030.0600 Neg Flow Verify: dP2 at -86,8 (Setpoint 90) [36m]	SLPM	-93,8 to -79,9	-89,4	✓	
40	0030.0610 Neg Flow Verify: dP2 at -82,0 (Setpoint 80) [36m]	SLPM	-88,8 to -75,2	-84,3	✓	
41	0030.0620 Neg Flow Verify: dP2 at -66,7 (Setpoint 70) [36m]	SLPM	-72,8 to -60,5	-69,8	✓	
42	0030.0630 Neg Flow Verify: dP2 at -63,8 (Setpoint 60) [36m]	SLPM	-69,9 to -57,8	-65,7	✓	
43	0030.0640 Neg Flow Verify: dP2 at -46,9 (Setpoint 50) [36m]	SLPM	-52,2 to -41,5	-48,7	✓	
44	0030.0650 Neg Flow Verify: dP2 at -43,5 (Setpoint 40) [36m]	SLPM	-48,7 to -38,3	-44,9	✓	
45	0030.0660 Neg Flow Verify: dP2 at -27,2 (Setpoint 30) [36m]	SLPM	-31,8 to -22,6	-28,4	✓	
46	0030.0670 Neg Flow Verify: dP2 at -19,7 (Setpoint 20) [36m]	SLPM	-24,7 to -14,7	-20,3	✓	
47	0030.0680 Neg Flow Verify: dP2 at -9,9 (Setpoint 10) [36m]	SLPM	-14,4 to -5,4	-9,9	✓	
48	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [36m]	SLPM	-8,8 to -1,2	-5,1	✓	
49	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [36m]	SLPM	-2,0 to 2,0	-0,1	✓	
50	0030.0950 Contr.Press.Ver.(8513): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,01	✓	
51	0030.0960 Monit.Press.Ver.(3799): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,03	✓	
52	0030.0160 Pprox Press. Ver.(9083): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
53	0030.0970 Contr.Press.Ver.(13247): P1 at 7,86 (SP 4, 1, 1)	cmH2O	6,86 to 8,86	7,86	✓	
54	0030.0980 Monit.Press.Ver.(9288): P2 at 7,86 (SP 4, 1, 1)	cmH2O	6,86 to 8,86	7,90	✓	
55	0030.0161 Pprox Press. Ver.(13844): P2 at 7,86 (SP 4, 1, 1)	cmH2O	6,86 to 8,86	7,89	✓	
56	0030.0990 Contr.Press.Ver.(19348): P1 at 18,04 (SP 18, 1, 1)	cmH2O	17,04 to 19,04	17,99	✓	
57	0030.1000 Monit.Press.Ver.(16353): P2 at 18,04 (SP 18, 1, 1)	cmH2O	17,04 to 19,04	18,03	✓	
58	0030.0162 Pprox Press. Ver.(19923): P2 at 18,04 (SP 18, 1, 1)	cmH2O	17,04 to 19,04	18,02	✓	
59	0030.1010 Contr.Press.Ver.(27793): P1 at 32,03 (SP 32, 1, 1)	cmH2O	31,03 to 33,03	32,00	✓	
60	0030.1020 Monit.Press.Ver.(26078): P2 at 32,03 (SP 32, 1, 1)	cmH2O	31,03 to 33,03	31,94	✓	
61	0030.0163 Pprox Press. Ver.(28310): P2 at 32,03 (SP 32, 1, 1)	cmH2O	31,03 to 33,03	32,03	✓	
62	0030.1030 Contr.Press.Ver.(36226): P1 at 45,98 (SP 46, 1, 1)	cmH2O	44,98 to 46,98	46,01	✓	
63	0030.1040 Monit.Press.Ver.(35786): P2 at 45,98 (SP 46, 1, 1)	cmH2O	44,98 to 46,98	45,90	✓	
64	0030.0164 Pprox Press. Ver.(36700): P2 at 45,98 (SP 46, 1, 1)	cmH2O	44,98 to 46,98	45,98	✓	
65	0030.1050 Contr.Press.Ver.(38661): P1 at 49,98 (SP 50, 1, 1)	cmH2O	48,98 to 50,98	49,98	✓	
66	0030.1060 Monit.Press.Ver.(38597): P2 at 49,98 (SP 50, 1, 1)	cmH2O	48,98 to 50,98	49,89	✓	
67	0030.0165 Pprox Press. Ver.(39123): P2 at 49,98 (SP 50, 1, 1)	cmH2O	48,98 to 50,98	50,05	✓	
68	0030.1070 Contr.Press.Ver.(47691): P1 at 64,99 (SP 65, 1, 1)	cmH2O	63,99 to 65,99	64,87	✓	
69	0030.1080 Monit.Press.Ver.(49035): P2 at 64,99 (SP 65, 1, 1)	cmH2O	63,99 to 65,99	64,96	✓	
70	0030.0166 Pprox Press. Ver.(48144): P2 at 64,99 (SP 65, 1, 1)	cmH2O	63,99 to 65,99	65,00	✓	

Test Started On 01/30/24 04:21:14Serial Number TV01406160EElapsed Test Time 42m 17sStatus **ERROR**

Page 2 of 3

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20