Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(35702): P1 at 45,51 (SP 46, 1,	cmH2O	44,51 to 46,51	46,00	✓	
72	0030.1040 Monit.Press.Ver.(36296): P2 at 45,51 (SP 46, 1,	cmH2O	44,51 to 46,51	45,48	✓	
73	0030.0164 Pprox Press. Ver.(35806): P2 at 45,51 (SP 46, 1	cmH2O	44,51 to 46,51	45,44	✓	
74	0030.1050 Contr.Press.Ver.(38105): P1 at 49,50 (SP 50, 1,	cmH2O	48,50 to 50,50	49,99	✓	
75	0030.1060 Monit.Press.Ver.(39084): P2 at 49,50 (SP 50, 1,	cmH2O	48,50 to 50,50	49,47	✓	
76	0030.0165 Pprox Press. Ver.(38185): P2 at 49,50 (SP 50, 1	cmH2O	48,50 to 50,50	49,45	✓	
77	0030.1070 Contr.Press.Ver.(47131): P1 at 64,39 (SP 65, 1,	cmH2O	63,39 to 65,39	65,00	✓	
78	0030.1080 Monit.Press.Ver.(49519): P2 at 64,39 (SP 65, 1,	cmH2O	63,39 to 65,39	64,37	✓	
79	0030.0166 Pprox Press. Ver.(47250): P2 at 64,39 (SP 65, 1	cmH2O	63,39 to 65,39	64,34	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,40	✓	
81	0030.1250 Airstream Temperature (Tair) at 26,00 [22m 53s]	C deg	21,00 to 41,00	30,89	✓	

Test Started On 04/24/24 02:57:29

Serial Number TV119042617 **Elapsed Test Time** 22m 53s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20