Test Software P/N 4102483

Tested Part Number 1054096



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,0 (Setpoint 75) [12m 37	ADC Cnts	LE 46448	45481	<u> </u>	_
37	0020.0580 Pos Flow Cal: dP2 at 87,9 (Setpoint 85) [12m 37	ADC Cnts	LE 47341	46450	-	—
38	0020.0590 Pos Flow Cal: dP2 at 97,0 (Setpoint 95) [12m 37	ADC Cnts	LE 48173	47342	\ \ \ \	
39	0020.0600 Pos Flow Cal: dP2 at 107,7 (Setpoint 105) [12m	ADC Cnts	LE 48914	48174	1	
40	0020.0610 Pos Flow Cal: dP2 at 117,8 (Setpoint 115) [12m	ADC Cnts	LE 49587	48915	1	_
41	0020.0620 Pos Flow Cal: dP2 at 128,2 (Setpoint 125) [12m	ADC Cnts	LE 50254	49588	1	
42	0020.0630 Pos Flow Cal: dP2 at 137,3 (Setpoint 135) [12m	ADC Cnts	LE 50823	50254	1	
43	0020.0640 Pos Flow Cal: dP2 at 147,8 (Setpoint 145) [12m	ADC Cnts	LE 52120	50824	1	
44	0020.0650 Pos Flow Cal: dP2 at 167,7 (Setpoint 170) [12m	ADC Cnts	LE 53361	52121	1	_
45	0020.0660 Pos Flow Cal: dP2 at 187,7 (Setpoint 190) [12m	ADC Cnts	LE 65435	53362	1	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [lnHg/ADC	EQ 0,000517	0,000517	√	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (inHg	1,91 to 4,09	2,52	√	
48	0020.0920 Write Sensors Cal Table [13m 2s]	N/A	Pass	TRUE	√	
49	0020.0970 Write Pprox Cal Table [14m 2s]	N/A	Pass	TRUE	√	
50	0020.0935 02 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	1	
51	0020.1030 Write Image Table [14m 23s]	N/A	Pass	TRUE	√	
						_
					\rightarrow	

Test Started On 02/26/24 12:19:07

Serial Number TV120051929 Elapsed Test Time 14m 23s

Status PASS

Page 2 **of** 2