Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## **Test Report**

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 02 Positive Flow Cal: dP1 at 64,2 (SP 65 @81,0P	ADC Cnts	LE 1847	1714	<b>√</b>	_
37	0021.0580 02 Positive Flow Cal: dP1 at 74,3 (SP 75 @80,4P	ADC Cnts	LE 1986	1848	<b>√</b>	
38	0021.0590 02 Positive Flow Cal: dP1 at 82,5 (SP 85 @79,7P	ADC Cnts	LE 2124	1987	<b>√</b>	_
39	0021.0600 02 Positive Flow Cal: dP1 at 93,8 (SP 95 @79,0P	ADC Cnts	LE 2270	2125	<b>√</b>	
40	0021.0610 02 Positive Flow Cal: dP1 at 105,5 (SP 105 @78,	ADC Cnts	LE 2377	2272	<b>√</b>	_
41	0021.0620 02 Positive Flow Cal: dP1 at 112,7 (SP 115 @77,	ADC Cnts	LE 2521	2378	<b>√</b>	
42	0021.0630 02 Positive Flow Cal: dP1 at 121,9 (SP 125 @76,	ADC Cnts	LE 2576	2522	<b>√</b>	
43	0021.0640 02 Positive Flow Cal: dP1 at 134,8 (SP 135 @75,	ADC Cnts	LE 2674	2577	<b>√</b>	_
44	0021.0650 O2 Positive Flow Cal: dP1 at 144,7 (SP 145 @74,	ADC Cnts	LE 2983	2675	<b>√</b>	-
45	0021.0660 O2 Positive Flow Cal: dP1 at 173,0 (SP 175 @71,	ADC Cnts	LE 3156	2984	<b>√</b>	
46	0021.0670 02 Positive Flow Cal: dP1 at 185,0 (SP 190 @68,	ADC Cnts	LE 3999	3157	<b>√</b>	
47	0021.0880 02 Sensor Cal (RANGE_FAILURE): 128 mV at ambien	N/A	Pass	FALSE		<u>×</u>
48	0021.0930 Write OBM Sensors Cal Table [17m 10s]	N/A	Pass	FALSE		<u>×</u>
						_
						_
						—
						—
					$\bot$	

**Test Started On** 02/07/24 11:16:30

Serial Number TV02001060E Elapsed Test Time 17m 10s

Page 2 **of** 2