Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 02 Positive Flow Cal: dPl at 64,7 (SP 65 @80,0P	ADC Cnts	LE 1913	1774	√
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,0 (SP 75 @79,2P	ADC Cnts	LE 2045	1914	✓
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,2 (SP 85 @78,5P)	ADC Cnts	LE 2180	2046	√
39	0021.0600 02 Positive Flow Cal: dP1 at 94,3 (SP 95 @77,5P	ADC Cnts	LE 2303	2181	√
40	0021.0610 02 Positive Flow Cal: dP1 at 105,3 (SP 105 @76,	ADC Cnts	LE 2402	2304	✓
41	0021.0620 O2 Positive Flow Cal: dPl at 114,3 (SP 115 @75,	ADC Cnts	LE 2503	2402	√
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,3 (SP 125 @74,	ADC Cnts	LE 2596	2504	✓
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,3 (SP 135 @74,	ADC Cnts	LE 2690	2597	✓
44	0021.0650 O2 Positive Flow Cal: dP1 at 145,2 (SP 145 @72,	ADC Cnts	LE 2943	2691	√
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,5 (SP 175 @70,	ADC Cnts	LE 3067	2944	√
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,5 (SP 190 @68,	ADC Cnts	LE 3999	3068	√
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 565 mV at ambient	N/A	Pass	TRUE	
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	→
					\bot

Test Started On 02/20/24 09:33:04

Serial Number TV017031585 Elapsed Test Time 14m 51s

Status PASS

Page 2 **of** 2