Test Software P/N 4102483

Tested Part Number 1040000

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA}} \ (\texttt{PM TEST})$



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,21	√	
72	0030.1250 Airstream Temperature (Tair) at 27,00 [45m 1s]	C deg	22,00 to 42,00	31,87	√	
						—
						_
						—
						—
						—
						—
			ļ		Ш	

Test Started On 01/31/24 11:19:03

Serial Number TV01406160E Elapsed Test Time 45m 1s

Status ERROR

Page 3 **of** 3