Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 64,0 (SP 65 @80,7P	ADC Cnts	LE 1958	1806	1/
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,0 (SP 75 @79,5P	ADC Cnts	LE 2111	1959	
38	0021.0590 O2 Positive Flow Cal: dP1 at 85,0 (SP 85 @78,1P	ADC Cnts	LE 2239	2112	<u> </u>
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,2 (SP 95 @76,8P	ADC Cnts	LE 2380	2240	<u> </u>
40	0021.0610 O2 Positive Flow Cal: dP1 at 105,0 (SP 105 @75,	ADC Cnts	LE 2472	2381	
41	0021.0620 O2 Positive Flow Cal: dP1 at 114,3 (SP 115 @73,	ADC Cnts	LE 2576	2473	─
42	0021.0630 O2 Positive Flow Cal: dP1 at 124,7 (SP 125 @72,	ADC Cnts	LE 2670	2578	-
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,0 (SP 135 @70,	ADC Cnts	LE 2767	2671	- √ -
44	0021.0650 O2 Positive Flow Cal: dP1 at 143,7 (SP 145 @69,	ADC Cnts	LE 3025	2768	─
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,0 (SP 175 @66,	ADC Cnts	LE 3150	3026	- √ -
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,1 (SP 190 @64,	ADC Cnts	LE 3999	3151	→
47	0021.0880 O2 Sensor Cal (NO_O2_ERROR): 592 mV at ambient	N/A	Pass	TRUE	- √ -
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	→
					+
					+
					+
					+
					- - -
					- - -
					++-
					++-
					+
					

Test Started On 02/27/24 05:17:22

Elapsed Test Time 17m 43s

Serial Number TV011062402

Status PASS

Page 2 **of** 2