PHILIPS
RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0040.0010 Keypad Backlight Test [5s]	N/A	Pass	TRUE	✓	
2	0040.0020 LCD Backlight Test [9s]	N/A	Pass	TRUE	✓	
3	0040.0030 Primary Audible Alarm Soft [17s]	N/A	Pass	TRUE	✓	
4	0040.0040 Primary Audible Alarm Soft heard by HW (1) [17s	N/A	Pass	TRUE	✓	
5	0040.0050 Primary Audible Alarm Loud [17s]	N/A	Pass	TRUE	✓	
6	0040.0060 Primary Audible Alarm Loud heard by HW (1) [17s	N/A	Pass	TRUE	✓	
7	0040.0070 Red Alarm LED [20s]	N/A	Pass	TRUE	✓	
8	0040.0080 Yellow Alarm LED [20s]	N/A	Pass	TRUE	✓	
9	0040.0090 Secondary Audible Alarm Soft [27s]	N/A	Pass	TRUE	✓	
10	0040.0100 Secondary Audible Alarm Soft heard by HW (1) [2	N/A	Pass	TRUE	✓	
11	0040.0110 Secondary Audible Alarm Loud [27s]	N/A	Pass	TRUE	✓	
12	0040.0120 Secondary Audible Alarm Loud heard by HW (1) [2	N/A	Pass	TRUE	✓	
13	0040.0130 LCD Brightness [35s]	N/A	Pass	TRUE	✓	
14	0040.0140 LCD screen test [41s]	N/A	Pass	TRUE	✓	

Test Started On 04/24/24 04:21:42

Serial Number TV019041206   **Elapsed Test Time** 41s

Status **PASS**