Test Software P/N 4102483 **Tested Part Number** 1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(36807): P1 at 45,86 (SP 46, 1,	cmH2O	44,86 to 46,86	45,99	✓	
72	0030.1040 Monit.Press.Ver.(34950): P2 at 45,86 (SP 46, 1,	cmH2O	44,86 to 46,86	45,86	✓	
73	0030.0164 Pprox Press. Ver.(36406): P2 at 45,86 (SP 46, 1	cmH2O	44,86 to 46,86	45,93	✓	
74	0030.1050 Contr.Press.Ver.(39199): P1 at 49,86 (SP 50, 1,	cmH2O	48,86 to 50,86	50,01	✓	
75	0030.1060 Monit.Press.Ver.(37709): P2 at 49,86 (SP 50, 1,	cmH2O	48,86 to 50,86	49,88	✓	
76	0030.0165 Pprox Press. Ver.(38832): P2 at 49,86 (SP 50, 1	cmH2O	48,86 to 50,86	49,98	✓	
77	0030.1070 Contr.Press.Ver.(47846): P1 at 63,29 (SP 65, 1,	cmH2O	62,29 to 64,29	64,50		✘
78	0030.1080 Monit.Press.Ver.(47756): P2 at 63,29 (SP 65, 1,	cmH2O	62,29 to 64,29	64,32		✘
79	0030.0166 Pprox Press. Ver.(47553): P2 at 63,29 (SP 65, 1	cmH2O	62,29 to 64,29	64,46		✘
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	27,99	✓	
81	0030.1250 Airstream Temperature (Tair) at 25,00 [24m 49s]	C deg	20,00 to 40,00	29,10	✓	

Test Started On 01/10/24 02:32:25

Serial Number TV114120511 **Elapsed Test Time** 24m 49s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20