Test Software P/N 4102483 **Tested Part Number** 1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(37128): P1 at 45,73 (SP 46, 1,	cmH2O	44,73 to 46,73	46,01	✓	
72	0030.1040 Monit.Press.Ver.(35437): P2 at 45,73 (SP 46, 1,	cmH2O	44,73 to 46,73	45,86	✓	
73	0030.0164 Pprox Press. Ver.(36707): P2 at 45,73 (SP 46, 1	cmH2O	44,73 to 46,73	45,82	✓	
74	0030.1050 Contr.Press.Ver.(39540): P1 at 49,73 (SP 50, 1,	cmH2O	48,73 to 50,73	49,99	✓	
75	0030.1060 Monit.Press.Ver.(38249): P2 at 49,73 (SP 50, 1,	cmH2O	48,73 to 50,73	49,88	✓	
76	0030.0165 Pprox Press. Ver.(39140): P2 at 49,73 (SP 50, 1	cmH2O	48,73 to 50,73	49,84	✓	
77	0030.1070 Contr.Press.Ver.(48091): P1 at 63,94 (SP 65, 1,	cmH2O	62,94 to 64,94	64,18	✓	
78	0030.1080 Monit.Press.Ver.(48209): P2 at 63,94 (SP 65, 1,	cmH2O	62,94 to 64,94	64,12	✓	
79	0030.0166 Pprox Press. Ver.(47680): P2 at 63,94 (SP 65, 1	cmH2O	62,94 to 64,94	64,13	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	27,93	✓	
81	0030.1250 Airstream Temperature (Tair) at 26,00 [20m 40s]	C deg	21,00 to 41,00	29,03	✓	

Test Started On 01/17/24 02:57:24

Serial Number TV110051733 **Elapsed Test Time** 20m 40s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20