Test Software P/N 4102483 **Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0020.0001 Ambient Air: RH=54,40%; T=23,60Cdeg [18s]	N/A	Pass	TRUE	✓	
2	0020.0010 O2 Sensor Heater On (not attempted for SW ver.1	N/A	None	None	✓	
3	0020.0010 Press Sensors Cal P1@3,98&50,06cmH2O/9889&34029	H2O/ADC C	,00151 to 0,0018	0,00167	✓	
4	0020.0020 Press Sensors Cal P1@3,98&50,06cmH2O/9889&34029	cmH2O	-18,22 to -9,05	-15,11	✓	
5	0020.0030 Press Sensors Cal P1@3,98&50,06cmH2O/9889&34029	ADC Cnts	5994 to 9998	9054	✓	
6	0020.0040 Press Sensors Cal P2@3,98&50,06cmH2O/9889&34029	H2O/ADC C	,00114 to 0,0018	0,00144	✓	
7	0020.0050 Press Sensors Cal P2@3,98&50,06cmH2O/9889&34029	cmH2O	-19,53 to 0,00	-5,64	✓	
8	0020.0060 Press Sensors Cal P2@3,98&50,06cmH2O/9889&34029	ADC Cnts	0 to 10399	3891	✓	
9	0020.0950 Press Sensors Cal Pprox@3,98&50,06cmH2O/9889&340	H2O/ADC C	,00151 to 0,0018	0,00167	✓	
10	0020.0960 Press Sensors Cal Pprox@3,98&50,06cmH2O/9889&34	cmH2O	-18,22 to -9,05	-14,69	✓	
11	0020.0970 Press Sensors Cal Pprox@3,98&50,06cmH2O/9889&34	ADC Cnts	5994 to 9998	8772	✓	
12	0020.0120 Neg Flow Cal: dP2 at -146,0 (Setpoint 145) [9m	ADC Cnts	GE 101	14332	✓	
13	0020.0130 Neg Flow Cal: dP2 at -132,7 (Setpoint 135) [9m	ADC Cnts	GE 14333	15288	✓	
14	0020.0140 Neg Flow Cal: dP2 at -122,7 (Setpoint 125) [9m	ADC Cnts	GE 15289	16079	✓	
15	0020.0150 Neg Flow Cal: dP2 at -113,5 (Setpoint 115) [9m	ADC Cnts	GE 16080	16895	✓	
16	0020.0160 Neg Flow Cal: dP2 at -103,3 (Setpoint 105) [9m	ADC Cnts	GE 16896	17783	✓	
17	0020.0170 Neg Flow Cal: dP2 at -93,5 (Setpoint 95) [9m 14	ADC Cnts	GE 17784	18799	✓	
18	0020.0180 Neg Flow Cal: dP2 at -83,5 (Setpoint 85) [9m 14	ADC Cnts	GE 18800	19945	✓	
19	0020.0190 Neg Flow Cal: dP2 at -73,0 (Setpoint 75) [9m 14	ADC Cnts	GE 19946	21008	✓	
20	0020.0200 Neg Flow Cal: dP2 at -64,0 (Setpoint 65) [9m 14	ADC Cnts	GE 21009	22058	✓	
21	0020.0210 Neg Flow Cal: dP2 at -53,5 (Setpoint 55) [9m 14	ADC Cnts	GE 22060	23350	✓	
22	0020.0220 Neg Flow Cal: dP2 at -43,3 (Setpoint 45) [9m 14	ADC Cnts	GE 23351	24725	✓	
23	0020.0230 Neg Flow Cal: dP2 at -33,8 (Setpoint 35) [9m 14	ADC Cnts	GE 24726	26209	✓	
24	0020.0240 Neg Flow Cal: dP2 at -23,0 (Setpoint 25) [9m 14	ADC Cnts	GE 26210	28059	✓	
25	0020.0250 Neg Flow Cal: dP2 at -15,0 (Setpoint 15) [9m 14	ADC Cnts	GE 28060	30179	✓	
26	0020.0260 Neg Flow Cal: dP2 at -5,0 (Setpoint 5) [9m 14s]	ADC Cnts	GE 30180	33261	✓	
27	0020.0270 Neg Flow Cal: dP2 at -0,0 (Setpoint 0) [9m 14s]	ADC Cnts	GE 33262	34500	✓	
28	0020.0450 Raw Zero Flow: dP2 at 0,0 (Setpoint 0) [9m 14s]	ADC Cnts	GE 0	34500	✓	
29	0020.0500 Pos Flow Cal: dP2 at 4,2 (Setpoint 5) [14m 11s]	ADC Cnts	LE 38260	35763	✓	
30	0020.0510 Pos Flow Cal: dP2 at 15,0 (Setpoint 15) [14m 11	ADC Cnts	LE 39455	38261	✓	
31	0020.0520 Pos Flow Cal: dP2 at 22,0 (Setpoint 25) [14m 11	ADC Cnts	LE 41746	39456	✓	
32	0020.0530 Pos Flow Cal: dP2 at 38,0 (Setpoint 35) [14m 11	ADC Cnts	LE 42781	41747	✓	
33	0020.0540 Pos Flow Cal: dP2 at 46,8 (Setpoint 45) [14m 11	ADC Cnts	LE 44024	42782	✓	
34	0020.0550 Pos Flow Cal: dP2 at 57,0 (Setpoint 55) [14m 11	ADC Cnts	LE 45243	44025	✓	
35	0020.0560 Pos Flow Cal: dP2 at 67,5 (Setpoint 65) [14m 11	ADC Cnts	LE 46237	45244	✓	

Test Started On 02/23/24 09:01:16

Serial Number TV015081807 **Elapsed Test Time** 21m 25s

Status **PASS**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20