Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
1	0040.0010 Keypad Backlight Test [5s]	N/A	Pass	TRUE	✓	
2	0040.0020 LCD Backlight Test [11s]	N/A	Pass	TRUE	✓	
3	0040.0030 Primary Audible Alarm Soft [3m 31s]	N/A	Pass	TRUE	✓	
4	0040.0040 Primary Audible Alarm Soft heard by HW (1) [3m	N/A	Pass	TRUE	✓	
5	0040.0050 Primary Audible Alarm Loud [3m 31s]	N/A	Pass	TRUE	✓	
6	0040.0060 Primary Audible Alarm Loud heard by HW (1) [3m	N/A	Pass	TRUE	✓	
7	0040.0070 Red Alarm LED [3m 36s]	N/A	Pass	TRUE	✓	
8	0040.0080 Yellow Alarm LED [3m 36s]	N/A	Pass	TRUE	✓	
9	0040.0090 Secondary Audible Alarm Soft [3m 48s]	N/A	Pass	TRUE	✓	
10	0040.0100 Secondary Audible Alarm Soft heard by HW (1) [3	N/A	Pass	TRUE	✓	
11	0040.0110 Secondary Audible Alarm Loud [3m 48s]	N/A	Pass	TRUE	✓	
12	0040.0120 Secondary Audible Alarm Loud heard by HW (1) [3	N/A	Pass	TRUE	✓	
13	0040.0130 LCD Brightness [3m 59s]	N/A	Pass	TRUE	✓	
14	0040.0140 LCD screen test [4m 0s]	N/A	Pass	TRUE	✓	

Test Started On 02/07/24 04:15:23

Serial Number TV113102301 **Elapsed Test Time** 4m 0s

Status **PASS**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Test Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20