Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Step	Description	Units	Limits	Results	Pass   Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 64,0 (SP 65 @76,7P	ADC Cnts	LE 1880	1748	1 / 1
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,3 (SP 75 @76,0P	ADC Cnts	LE 2015	1881	<del> </del>
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,3 (SP 85 @75,1P	ADC Cnts	LE 2132	2016	<u> </u>
39	0021.0600 02 Positive Flow Cal: dP1 at 94,9 (SP 95 @74,2P	ADC Cnts	LE 2254	2133	<u> </u>
40	0021.0610 02 Positive Flow Cal: dP1 at 104,8 (SP 105 @73,	ADC Cnts	LE 2365	2255	<u> </u>
41	0021.0620 O2 Positive Flow Cal: dP1 at 115,0 (SP 115 @72,	ADC Cnts	LE 2467	2366	<del>-                                     </del>
42	0021.0630 O2 Positive Flow Cal: dP1 at 125,3 (SP 125 @71,	ADC Cnts	LE 2561	2468	<del>-                                     </del>
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,8 (SP 135 @70,	ADC Cnts	LE 2646	2562	<del>-                                     </del>
44	0021.0650 O2 Positive Flow Cal: dP1 at 144,5 (SP 145 @69,	ADC Cnts	LE 2910	2646	<b>│</b>
45	0021.0660 O2 Positive Flow Cal: dP1 at 175,2 (SP 175 @65,	ADC Cnts	LE 3034	2910	<del>-   √   -  </del>
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,8 (SP 190 @63,	ADC Cnts	LE 3999	3035	<del>-   √   -  </del>
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 586 mV at ambient	N/A	Pass	TRUE	<b>│</b>
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	<del>-   √   -  </del>
					+
					+
					<del>                                     </del>
					<del>                                     </del>
					<del>       </del>
					<del>       </del>
					<del>       </del>
					<del>                                     </del>
				ļ	$-\!\!\!\!-\!\!\!\!\!-\!\!\!\!\!-$

**Test Started On** 02/12/24 02:44:41

Elapsed Test Time 15m 16s

Serial Number TV014061613

Status PASS

Page 2 **of** 2