Test Software P/N 4102483 Tested Part Number 1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 Prepared By P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(36556): P1 at 45,92 (SP 46, 1,	cmH2O	44,92 to 46,92	45,99	✓	
72	0030.1040 Monit.Press.Ver.(35638): P2 at 45,92 (SP 46, 1,	cmH2O	44,92 to 46,92	45,99	✓	
73	0030.0164 Pprox Press. Ver.(36926): P2 at 45,92 (SP 46, 1	cmH2O	44,92 to 46,92	46,13	✓	
74	0030.1050 Contr.Press.Ver.(38969): P1 at 49,87 (SP 50, 1,	cmH2O	48,87 to 50,87	49,99	✓	
75	0030.1060 Monit.Press.Ver.(38419): P2 at 49,87 (SP 50, 1,	cmH2O	48,87 to 50,87	49,95	✓	
76	0030.0165 Pprox Press. Ver.(39322): P2 at 49,87 (SP 50, 1	cmH2O	48,87 to 50,87	50,12	✓	
77	0030.1070 Contr.Press.Ver.(47933): P1 at 64,73 (SP 65, 1,	cmH2O	63,73 to 65,73	64,85	✓	
78	0030.1080 Monit.Press.Ver.(48760): P2 at 64,73 (SP 65, 1,	cmH2O	63,73 to 65,73	64,73	✓	
79	0030.0166 Pprox Press. Ver.(48217): P2 at 64,73 (SP 65, 1	cmH2O	63,73 to 65,73	64,79	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,07	✓	
81	[22m 56s]			TERMINATED		✘
82	0030.1250 Airstream Temperature (Tair) at 26,00 [23m 26s]	C deg	21,00 to 41,00	28,98	✓	

Test Started On 01/08/24 02:43:36

Serial Number TV110052651 Elapsed Test Time 23m 26s

Status **TERMINATED**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 1FF8, 05/05/23 08:20