Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (REPAIR TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,8 (Setpoint 75) [12m 38	ADC Cnts	LE 46674	45708	✓	
37	0020.0580 Pos Flow Cal: dP2 at 87,0 (Setpoint 85) [12m 38	ADC Cnts	LE 47714	46675	✓	
38	0020.0590 Pos Flow Cal: dP2 at 97,0 (Setpoint 95) [12m 38	ADC Cnts	LE 48618	47716	✓	
39	0020.0600 Pos Flow Cal: dP2 at 107,3 (Setpoint 105) [12m	ADC Cnts	LE 49351	48619	✓	
40	0020.0610 Pos Flow Cal: dP2 at 117,2 (Setpoint 115) [12m	ADC Cnts	LE 50089	49352	✓	
41	0020.0620 Pos Flow Cal: dP2 at 128,0 (Setpoint 125) [12m	ADC Cnts	LE 50660	50090	✓	
42	0020.0630 Pos Flow Cal: dP2 at 136,5 (Setpoint 135) [12m	ADC Cnts	LE 51416	50661	✓	
43	0020.0640 Pos Flow Cal: dP2 at 148,3 (Setpoint 145) [12m	ADC Cnts	LE 52604	51417	✓	
44	0020.0650 Pos Flow Cal: dP2 at 169,2 (Setpoint 170) [12m	ADC Cnts	LE 53698	52605	✓	
45	0020.0660 Pos Flow Cal: dP2 at 188,2 (Setpoint 190) [12m	ADC Cnts	LE 65435	53699	✓	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [	inHg/ADC	EQ 0,000517	0,000517	✓	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (	inHg	1,91 to 4,09	2,50	✓	
48	0020.0920 Write Sensors Cal Table [13m 4s]	N/A	Pass	TRUE	✓	
49	0020.0970 Write Pprox Cal Table [13m 19s]	N/A	Pass	TRUE	✓	
50	0020.0935 O2 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	✓	
51	0020.1030 Write Image Table [13m 40s]	N/A	Pass	TRUE	✓	

Test Started On 03/27/24 12:05:16

Serial Number TV112010957 **Elapsed Test Time** 13m 40s

Status **PASS**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 77D6, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20