PHILIPS
RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
176	0050.0180 Device's SD Card Inserted: SD Detected (3) [2 h:	N/A	EQ YES	YES	✓	
177	0050.0190 Device's SD Card Inserted: SD Recognized (3) [2	N/A	EQ YES	YES	✓	
178	0050.0195 Device's SD Card Inserted: SD Unlocked (3) [2 h:	N/A	EQ YES	YES	✓	
179	0050.0200 Significant Event Log (198 entries): [2 hr 8m 2:	N/A	Pass	TRUE	✓	
180	0050.0210 Engineering Debug Log (111 entries): [2 hr 9m	N/A	Pass	TRUE	✓	
181	0050.0020 Reset Operational Hours (4189,7 Hrs) [2 hr 9m 3	N/A	Pass	TRUE	✓	

Test Started On 01/16/24 02:52:59

Serial Number TV116080284   **Elapsed Test Time** 2 hr 9m 39s

Status **PASS**