

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0280 Pos Flow Verify: dP2 at 48,0 (Setpoint 50) [11m	SLPM	42,6 to 53,4	46,9	✓	
37	0030.0290 Pos Flow Verify: dP2 at 37,3 (Setpoint 40) [11m	SLPM	32,3 to 42,3	36,6	✓	
38	0030.0300 Pos Flow Verify: dP2 at 27,8 (Setpoint 30) [11m	SLPM	23,2 to 32,4	27,2	✓	
39	0030.0310 Pos Flow Verify: dP2 at 19,8 (Setpoint 20) [11m	SLPM	14,8 to 24,8	19,4	✓	
40	0030.0320 Pos Flow Verify: dP2 at 10,0 (Setpoint 10) [11m	SLPM	5,5 to 14,5	9,9	✓	
41	0030.0330 Pos Flow Verify: dP2 at 5,0 (Setpoint 5) [11m 1	SLPM	1,2 to 8,8	4,7	✓	
42	0030.0510 Zero Flow Verify: dP2 at 0,0 (Setpoint 0) [11m	SLPM	-2,0 to 2,0	-0,0	✓	
43	0030.0550 Neg Flow Verify: dP2 at -138,7 (Setpoint 140) [SLPM	-147,7 to -129,6	-132,5	✓	
44	0030.0560 Neg Flow Verify: dP2 at -128,0 (Setpoint 130) [SLPM	-136,6 to -119,4	-122,3	✓	
45	0030.0570 Neg Flow Verify: dP2 at -117,8 (Setpoint 120) [SLPM	-126,0 to -109,6	-112,8	✓	
46	0030.0580 Neg Flow Verify: dP2 at -108,0 (Setpoint 110) [SLPM	-115,8 to -100,2	-104,6	✓	
47	0030.0590 Neg Flow Verify: dP2 at -98,7 (Setpoint 100) [1	SLPM	-106,1 to -91,2	-94,9	✓	
48	0030.0600 Neg Flow Verify: dP2 at -88,8 (Setpoint 90) [16	SLPM	-95,9 to -81,8	-85,0	✓	
49	0030.0610 Neg Flow Verify: dP2 at -77,7 (Setpoint 80) [16	SLPM	-84,3 to -71,1	-74,3	✓	
50	0030.0620 Neg Flow Verify: dP2 at -67,2 (Setpoint 70) [16	SLPM	-73,4 to -61,0	-65,7	✓	
51	0030.0630 Neg Flow Verify: dP2 at -57,3 (Setpoint 60) [16	SLPM	-63,1 to -51,5	-55,7	✓	
52	0030.0640 Neg Flow Verify: dP2 at -48,9 (Setpoint 50) [16	SLPM	-54,3 to -43,4	-46,7	✓	
53	0030.0650 Neg Flow Verify: dP2 at -38,0 (Setpoint 40) [16	SLPM	-43,0 to -33,0	-36,4	✓	
54	0030.0660 Neg Flow Verify: dP2 at -28,0 (Setpoint 30) [16	SLPM	-32,6 to -23,4	-27,6	✓	
55	0030.0670 Neg Flow Verify: dP2 at -19,0 (Setpoint 20) [16	SLPM	-24,0 to -14,0	-19,4	✓	
56	0030.0680 Neg Flow Verify: dP2 at -9,7 (Setpoint 10) [16m	SLPM	-14,2 to -5,2	-9,4	✓	
57	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [16m	SLPM	-8,8 to -1,2	-4,8	✓	
58	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [16m	SLPM	-2,0 to 2,0	-0,0	✓	
59	0030.0950 Contr.Press.Ver.(8082): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,04	✓	
60	0030.0960 Monit.Press.Ver.(4132): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
61	0030.0160 Pprox Press. Ver.(8050): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,08	✓	
62	0030.0970 Contr.Press.Ver.(10471): P1 at 3,92 (SP 4, 1, 1)	cmH2O	2,92 to 4,92	3,99	✓	
63	0030.0980 Monit.Press.Ver.(6876): P2 at 3,92 (SP 4, 1, 1)	cmH2O	2,92 to 4,92	3,75	✓	
64	0030.0161 Pprox Press. Ver.(10416): P2 at 3,92 (SP 4, 1,	cmH2O	2,92 to 4,92	4,01	✓	
65	0030.0990 Contr.Press.Ver.(18926): P1 at 17,95 (SP 18, 1,	cmH2O	16,95 to 18,95	17,99	✓	
66	0030.1000 Monit.Press.Ver.(16677): P2 at 17,95 (SP 18, 1,	cmH2O	16,95 to 18,95	17,76	✓	
67	0030.0162 Pprox Press. Ver.(18892): P2 at 17,95 (SP 18, 1	cmH2O	16,95 to 18,95	18,00	✓	
68	0030.1010 Contr.Press.Ver.(27375): P1 at 31,91 (SP 32, 1,	cmH2O	30,91 to 32,91	32,00	✓	
69	0030.1020 Monit.Press.Ver.(26436): P2 at 31,91 (SP 32, 1,	cmH2O	30,91 to 32,91	31,73	✓	
70	0030.0163 Pprox Press. Ver.(27289): P2 at 31,91 (SP 32, 1	cmH2O	30,91 to 32,91	31,96	✓	

Test Started On 01/04/24 02:55:24Serial Number TV11909263DElapsed Test Time 59m 54sStatus **PASS**

Page 2 of 6

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 1FF8, 05/05/23 08:20