Test Software P/N 4102483 Tested Part Number LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 Prepared By P. Pascal

PHILIPS RESPIRONICS

# Test Report

| Step | Description | Units | Limits | Results | Pass | Fail |
|---|---|---|---|---|---|---|
| 141 | 0040.0700 SD Locked Card Inserted: SD Detected (7) [41m 3 | N/A | EQ YES | YES | ✓ | |
| 142 | 0040.0710 SD Locked Card Inserted: SD Recognized (7) [41m | N/A | EQ YES | YES | ✓ | |
| 143 | 0040.0720 SD Locked Card Inserted: Write Protected Mode ( | N/A | EQ YES | YES | ✓ | |
| 144 | 0040.0730 SD Card Removed: SD Detected (4) [42m 17s] | N/A | EQ NO | NO | ✓ | |
| 145 | 0040.0740 SD Card Removed: SD Recognized (4) [42m 17s] | N/A | EQ NO | NO | ✓ | |
| 146 | 0040.0750 SD Card Removed: Write Protected Mode (4) [42m | N/A | EQ YES | YES | ✓ | |
| 147 | 0040.0780 SD Unlocked Card Inserted: Write Mode Enable (3 | N/A | Pass | TRUE | ✓ | |
| 148 | 0040.0790 SD: File Write [42m 30s] | N/A | Pass | TRUE | ✓ | |
| 149 | 0040.0800 SD: File Length [42m 31s] | N/A | Pass | TRUE | ✓ | |
| 150 | 0040.0810 SD: File Read [42m 32s] | N/A | Pass | TRUE | ✓ | |
| 151 | 0040.0820 SD: File Erase [42m 33s] | N/A | Pass | TRUE | ✓ | |
| 152 | 0040.0840 Nurse Call OFF (NO: 9999,00KOhm, NC: 0,00Ohm) [ | N/A | Pass | TRUE | ✓ | |
| 153 | 0040.0850 Nurse Call ON (NO: 0,00Ohm, NC: 9999,00KOhm) [4 | N/A | Pass | TRUE | ✓ | |
| 154 | 0040.0900 P1 with Auto Null Valve Closed at 30,02cmH2O [4 | cmH2O | 27,02 to 33,02 | 29,99 | ✓ | |
| 155 | 0040.0910 P1 with Auto Null Valve Opened at 64,90cmH2O [4 | cmH2O | -3,00 to 3,00 | 0,03 | ✓ | |
| 156 | 0040.0950 Cooling Fans: Stirring Fan [43m 25s] | N/A | Pass | TRUE | ✓ | |
| 157 | 0040.0960 Cooling Fans: Battery Fan [43m 27s] | N/A | Pass | TRUE | ✓ | |
| 158 | 0040.0970 Cooling Fans: Exhaust Fan [43m 45s] | N/A | Pass | TRUE | ✓ | |
| 159 | 0040.1300 Current Consumption in Sleep Mode (1) [45m 44s] | A | LE 0,20 | 0,00 | ✓ | |
| 160 | 0040.1400 O2 Low Flow @ 20,20PSIG [48m 9s] | SLPM | GE 7 | 11 | ✓ | |
| 161 | 0040.1410 DSP CRC: JTAG ID Code [48m 50s] | N/A | EQ 01F4001D | 01F4001D | ✓ | |
| 162 | 0040.1420 DSP CRC: DSP Program CRC [48m 50s] | N/A | EQ FE0F | FE0F | ✓ | |

Test Started On 06/06/24 09:37:49

Serial Number TV11909271C

Elapsed Test Time 48m 50s

Status **FAIL**



Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20