

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0030.0570 Neg Flow Verify: dP2 at -117,0 (Setpoint 120) [41m	SLPM	-125,2 to -108,8	-122,0	✓	
37	0030.0580 Neg Flow Verify: dP2 at -107,7 (Setpoint 110) [41m	SLPM	-115,5 to -99,9	-111,7	✓	
38	0030.0590 Neg Flow Verify: dP2 at -97,0 (Setpoint 100) [41m	SLPM	-104,4 to -89,6	-100,3	✓	
39	0030.0600 Neg Flow Verify: dP2 at -86,7 (Setpoint 90) [41m	SLPM	-93,6 to -79,7	-89,8	✓	
40	0030.0610 Neg Flow Verify: dP2 at -76,0 (Setpoint 80) [41m	SLPM	-82,5 to -69,5	-79,3	✓	
41	0030.0620 Neg Flow Verify: dP2 at -66,0 (Setpoint 70) [41m	SLPM	-72,1 to -59,9	-68,8	✓	
42	0030.0630 Neg Flow Verify: dP2 at -65,0 (Setpoint 60) [41m	SLPM	-71,1 to -58,9	-66,4	✓	
43	0030.0640 Neg Flow Verify: dP2 at -46,3 (Setpoint 50) [41m	SLPM	-51,7 to -41,0	-48,0	✓	
44	0030.0650 Neg Flow Verify: dP2 at -43,0 (Setpoint 40) [41m	SLPM	-48,2 to -37,8	-44,5	✓	
45	0030.0660 Neg Flow Verify: dP2 at -33,2 (Setpoint 30) [41m	SLPM	-38,0 to -28,3	-34,1	✓	
46	0030.0670 Neg Flow Verify: dP2 at -19,5 (Setpoint 20) [41m	SLPM	-24,5 to -14,5	-20,5	✓	
47	0030.0680 Neg Flow Verify: dP2 at -9,5 (Setpoint 10) [41m	SLPM	-14,0 to -5,0	-9,8	✓	
48	0030.0690 Neg Flow Verify: dP2 at -5,0 (Setpoint 5) [41m	SLPM	-8,8 to -1,2	-5,0	✓	
49	0030.0700 Neg Flow Verify: dP2 at -0,0 (Setpoint 0) [41m	SLPM	-2,0 to 2,0	-0,1	✓	
50	0030.0950 Contr.Press.Ver.(8550): P1 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
51	0030.0960 Monit.Press.Ver.(3718): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
52	0030.0160 Pprox Press. Ver.(9082): P2 at 0,00 (SP 0, 1, 1)	cmH2O	-1,00 to 1,00	0,00	✓	
53	0030.0970 Contr.Press.Ver.(13321): P1 at 7,98 (SP 4, 1, 1)	cmH2O	6,98 to 8,98	7,88	✓	
54	0030.0980 Monit.Press.Ver.(9242): P2 at 7,98 (SP 4, 1, 1)	cmH2O	6,98 to 8,98	7,82	✓	
55	0030.0161 Pprox Press. Ver.(13876): P2 at 7,98 (SP 4, 1, 1)	cmH2O	6,98 to 8,98	7,95	✓	
56	0030.0990 Contr.Press.Ver.(19412): P1 at 18,16 (SP 18, 1, 1)	cmH2O	17,16 to 19,16	18,00	✓	
57	0030.1000 Monit.Press.Ver.(16317): P2 at 18,16 (SP 18, 1, 1)	cmH2O	17,16 to 19,16	17,95	✓	
58	0030.0162 Pprox Press. Ver.(19971): P2 at 18,16 (SP 18, 1, 1)	cmH2O	17,16 to 19,16	18,10	✓	
59	0030.1010 Contr.Press.Ver.(27844): P1 at 32,16 (SP 32, 1, 1)	cmH2O	31,16 to 33,16	31,99	✓	
60	0030.1020 Monit.Press.Ver.(26053): P2 at 32,16 (SP 32, 1, 1)	cmH2O	31,16 to 33,16	31,92	✓	
61	0030.0163 Pprox Press. Ver.(28366): P2 at 32,16 (SP 32, 1, 1)	cmH2O	31,16 to 33,16	32,10	✓	
62	0030.1030 Contr.Press.Ver.(36266): P1 at 46,14 (SP 46, 1, 1)	cmH2O	45,14 to 47,14	45,96	✓	
63	0030.1040 Monit.Press.Ver.(35755): P2 at 46,14 (SP 46, 1, 1)	cmH2O	45,14 to 47,14	45,88	✓	
64	0030.0164 Pprox Press. Ver.(36788): P2 at 46,14 (SP 46, 1, 1)	cmH2O	45,14 to 47,14	46,10	✓	
65	0030.1050 Contr.Press.Ver.(38690): P1 at 50,11 (SP 50, 1, 1)	cmH2O	49,11 to 51,11	50,00	✓	
66	0030.1060 Monit.Press.Ver.(38562): P2 at 50,11 (SP 50, 1, 1)	cmH2O	49,11 to 51,11	49,87	✓	
67	0030.0165 Pprox Press. Ver.(39139): P2 at 50,11 (SP 50, 1, 1)	cmH2O	49,11 to 51,11	50,00	✓	
68	0030.1070 Contr.Press.Ver.(47734): P1 at 65,10 (SP 65, 1, 1)	cmH2O	64,10 to 66,10	65,01	✓	
69	0030.1080 Monit.Press.Ver.(49003): P2 at 65,10 (SP 65, 1, 1)	cmH2O	64,10 to 66,10	64,75	✓	
70	0030.0166 Pprox Press. Ver.(48179): P2 at 65,10 (SP 65, 1, 1)	cmH2O	64,10 to 66,10	65,15	✓	

Test Started On 01/31/24 11:19:03Serial Number TV01406160EElapsed Test Time 45m 1sStatus **ERROR**

Page 2 of 3

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Troubleshooting: Group Level mode.
 File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
 Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20