Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0021.0570 O2 Positive Flow Cal: dP1 at 65,0 (SP 65 @80,8P	ADC Cnts	LE 1922	1783	√	
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,2 (SP 75 @79,8P	ADC Cnts	LE 2062	1923	√	
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,3 (SP 85 @78,6P	ADC Cnts	LE 2180	2063	√	
39	0021.0600 O2 Positive Flow Cal: dP1 at 93,5 (SP 95 @77,4P	ADC Cnts	LE 2294	2181	√	
40	0021.0610 O2 Positive Flow Cal: dP1 at 103,7 (SP 105 @76,	ADC Cnts	LE 2402	2295	√	
41	0021.0620 O2 Positive Flow Cal: dP1 at 114,3 (SP 115 @74,	ADC Cnts	LE 2509	2403	√	
42	0021.0630 O2 Positive Flow Cal: dPl at 124,0 (SP 125 @73,	ADC Cnts	LE 2608	2510	√	
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,7 (SP 135 @71,	ADC Cnts	LE 2705	2609	√	
44	0021.0650 O2 Positive Flow Cal: dP1 at 145,0 (SP 145 @69,	ADC Cnts	LE 2974	2706	√	
45	0021.0660 O2 Positive Flow Cal: dP1 at 175,0 (SP 175 @66,	ADC Cnts	LE 3096	2974	√	
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,5 (SP 190 @64,	ADC Cnts	LE 3999	3097	√	
47	0021.0880 02 Sensor Cal (RANGE_FAILURE): 3 mV at ambient	N/A	Pass	FALSE		×
48	0021.0930 Write OBM Sensors Cal Table [17m 51s]	N/A	Pass	FALSE		×
						$\overline{}$
						-
						\vdash
						<u> </u>
						
						\vdash
						<u> </u>
						<u> </u>
						<u> </u>
						-
						
						<u> </u>
					\perp	<u> </u>
						<u> </u>
						<u> </u>

Test Started On 02/02/24 03:04:04

Serial Number TV014061613 Elapsed Test Time 17m 51s

Status FAIL

Page 2 **of** 2