Test Software P/N 4102483

**Tested Part Number** 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



**Software Revision** 24.0.0.0

Prepared By P. Pascal

## Test Report

Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 O2 Positive Flow Cal: dPl at 64,5 (SP 65 @81,7P	ADC Cnts	LE 1997	1855	<b>√</b>
37	0021.0580 O2 Positive Flow Cal: dP1 at 74,3 (SP 75 @81,0P	ADC Cnts	LE 2138	1998	✓
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,7 (SP 85 @80,2P	ADC Cnts	LE 2266	2139	<b>√</b>
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,7 (SP 95 @79,3P	ADC Cnts	LE 2393	2267	<b>√</b>
40	0021.0610 02 Positive Flow Cal: dP1 at 104,2 (SP 105 @78,	ADC Cnts	LE 2506	2394	✓
41	0021.0620 O2 Positive Flow Cal: dPl at 114,0 (SP 115 @77,	ADC Cnts	LE 2608	2507	<b>√</b>
42	0021.0630 O2 Positive Flow Cal: dP1 at 123,7 (SP 125 @76,	ADC Cnts	LE 2701	2609	✓
43	0021.0640 O2 Positive Flow Cal: dP1 at 134,2 (SP 135 @75,	ADC Cnts	LE 2792	2702	✓
44	0021.0650 O2 Positive Flow Cal: dP1 at 143,1 (SP 145 @74,	ADC Cnts	LE 3052	2793	<b>√</b>
45	0021.0660 O2 Positive Flow Cal: dP1 at 174,3 (SP 175 @71,	ADC Cnts	LE 3175	3053	<b>√</b>
46	0021.0670 O2 Positive Flow Cal: dP1 at 189,2 (SP 190 @69,	ADC Cnts	LE 3999	3176	<b>√</b>
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 537 mV at ambient	N/A	Pass	TRUE	<b> </b>
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	<b>→</b>
					$\bot$

**Test Started On** 02/14/24 02:37:27

Serial Number TV012043013 Elapsed Test Time 18m 44s

Status PASS

Page 2 **of** 2