Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
141	0040.0700 SD Locked Card Inserted: SD Detected (7) [27m 3	N/A	EQ YES	YES	✓	
142	0040.0710 SD Locked Card Inserted: SD Recognized (7) [27m	N/A	EQ YES	YES	✓	
143	0040.0720 SD Locked Card Inserted: Write Protected Mode (	N/A	EQ YES	YES	✓	
144	0040.0730 SD Card Removed: SD Detected (4) [27m 45s]	N/A	EQ NO	NO	✓	
145	0040.0740 SD Card Removed: SD Recognized (4) [27m 45s]	N/A	EQ NO	NO	✓	
146	0040.0750 SD Card Removed: Write Protected Mode (4) [27m	N/A	EQ YES	YES	✓	
147	0040.0780 SD Unlocked Card Inserted: Write Mode Enable (3	N/A	Pass	TRUE	✓	
148	0040.0790 SD: File Write [28m 6s]	N/A	Pass	TRUE	✓	
149	0040.0800 SD: File Length [28m 7s]	N/A	Pass	TRUE	✓	
150	0040.0810 SD: File Read [28m 8s]	N/A	Pass	TRUE	✓	
151	0040.0820 SD: File Erase [28m 9s]	N/A	Pass	TRUE	✓	
152	0040.0840 Nurse Call OFF (NO: 9999,00KOhm, NC: 1,00Ohm) [	N/A	Pass	TRUE	✓	
153	0040.0850 Nurse Call ON (NO: 1,00Ohm, NC: 9999,00KOhm) [2	N/A	Pass	TRUE	✓	
154	0040.0900 P1 with Auto Null Valve Closed at 30,04cmH2O [2	cmH2O	27,04 to 33,04	29,98	✓	
155	0040.0910 P1 with Auto Null Valve Opened at 61,98cmH2O [2	cmH2O	-3,00 to 3,00	0,00	✓	
156	0040.0950 Cooling Fans: Stirring Fan [28m 58s]	N/A	Pass	FALSE		✘
157	0040.0960 Cooling Fans: Battery Fan [29m 0s]	N/A	Pass	TRUE	✓	
158	0040.0970 Cooling Fans: Exhaust Fan [29m 16s]	N/A	Pass	TRUE	✓	
159	0040.1300 Current Consumption in Sleep Mode (1) [30m 29s]	A	LE 0,20	0,00	✓	
160	0040.1400 O2 Low Flow @ 0,00PSIG [31m 50s]	SLPM		ERROR 8025		✘
161	0040.1410 DSP CRC: JTAG ID Code [31m 51s]	N/A	EQ 01F4001D	01F4001D	✓	
162	0040.1420 DSP CRC: DSP Program CRC [31m 51s]	N/A	EQ FE0F	FE0F	✓	
163	[31m 52s]			TERMINATED		✘
164	[31m 53s]			TERMINATED		✘
165	[31m 59s]			TERMINATED		✘

Test Started On 02/16/24 12:49:01

Serial Number TV114121231 **Elapsed Test Time** 31m 59s

Status **TERMINATED**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20