Test Software P/N 4102483 **Tested Part Number** LA1054096

Test Software Name Trilogy FSA (PM TEST)

Software Revision 24.0.0.0 **Prepared By** P. Pascal

PHILIPS RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
71	0030.1030 Contr.Press.Ver.(36169): P1 at 46,12 (SP 46, 1,	cmH2O	45,12 to 47,12	45,99	✓	
72	0030.1040 Monit.Press.Ver.(36843): P2 at 46,12 (SP 46, 1,	cmH2O	45,12 to 47,12	45,76	✓	
73	0030.0164 Pprox Press. Ver.(35926): P2 at 46,12 (SP 46, 1	cmH2O	45,12 to 47,12	45,75	✓	
74	0030.1050 Contr.Press.Ver.(38626): P1 at 50,16 (SP 50, 1,	cmH2O	49,16 to 51,16	49,98	✓	
75	0030.1060 Monit.Press.Ver.(39697): P2 at 50,16 (SP 50, 1,	cmH2O	49,16 to 51,16	49,77	✓	
76	0030.0165 Pprox Press. Ver.(38350): P2 at 50,16 (SP 50, 1	cmH2O	49,16 to 51,16	49,79	✓	
77	0030.1070 Contr.Press.Ver.(47215): P1 at 64,27 (SP 65, 1,	cmH2O	63,27 to 65,27	63,97	✓	
78	0030.1080 Monit.Press.Ver.(49598): P2 at 64,27 (SP 65, 1,	cmH2O	63,27 to 65,27	63,82	✓	
79	0030.0166 Pprox Press. Ver.(46938): P2 at 64,27 (SP 65, 1	cmH2O	63,27 to 65,27	63,83	✓	
80	0030.1150 Atmospheric Pressure Verify: Patm at 28,05,28,0	inHg	27,45 to 28,65	28,28	✓	
81	0030.1250 Airstream Temperature (Tair) at 26,00 [19m 26s]	C deg	21,00 to 41,00	30,74	✓	

Test Started On 08/27/24 03:36:21

Serial Number TV119091016 **Elapsed Test Time** 19m 26s

Status **FAIL**

Audit Info Printed to:doPDF v7 from RESPIRONICS-PC in Production mode.
File: C:\Program Files\Trilogy Service Cal And Test\Trilogy Repair_PM Cal And Test.exe, Size: 8674304, 05/09/23 10:38
Station Cfg CRC: 5219, 09/01/21 01:39, Test Cfg CRC: 94CF, 09/01/21 01:39, Models Cfg CRC: 2F86, 05/05/23 08:20