Test Software P/N 4102483

Tested Part Number LA1054096

 $\textbf{Test Software Name} \ \underline{\texttt{Trilogy FSA (REPAIR TEST)}}$

PHILIPS RESPIRONICS

Software Revision 24.0.0.0 Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,7 (Setpoint 75) [14m 19	ADC Cnts	LE 47268	46368	<u> </u>	
37	0020.0580 Pos Flow Cal: dP2 at 87,3 (Setpoint 85) [14m 19	ADC Cnts	LE 48232	47269	1	
38	0020.0590 Pos Flow Cal: dP2 at 96,7 (Setpoint 95) [14m 19	ADC Cnts	LE 49035	48233	1	
39	0020.0600 Pos Flow Cal: dP2 at 106,7 (Setpoint 105) [14m	ADC Cnts	LE 49884	49036	1	
40	0020.0610 Pos Flow Cal: dP2 at 117,9 (Setpoint 115) [14m	ADC Cnts	LE 50568	49885	√	
41	0020.0620 Pos Flow Cal: dP2 at 127,2 (Setpoint 125) [14m	ADC Cnts	LE 51113	50569	1	
42	0020.0630 Pos Flow Cal: dP2 at 134,1 (Setpoint 135) [14m	ADC Cnts	LE 51672	51114	√	
43	0020.0640 Pos Flow Cal: dP2 at 142,3 (Setpoint 145) [14m	ADC Cnts	LE 53125	51673	√	
44	0020.0650 Pos Flow Cal: dP2 at 168,0 (Setpoint 170) [14m	ADC Cnts	LE 54372	53126	√	
45	0020.0660 Pos Flow Cal: dP2 at 188,5 (Setpoint 190) [14m	ADC Cnts	LE 65435	54373	√	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [lnHg/ADC	EQ 0,000517	0,000517	√	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (inHg	1,91 to 4,09	2,80	√	
48	0020.0920 Write Sensors Cal Table [14m 43s]	N/A	Pass	TRUE	√	
49	0020.0970 Write Pprox Cal Table [14m 59s]	N/A	Pass	TRUE	√	
50	0020.0935 02 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	√	
51	0020.1030 Write Image Table [15m 20s]	N/A	Pass	TRUE	√	

Test Started On 03/25/24 08:41:07

Serial Number TV118062688 Elapsed Test Time 15m 20s

Status PASS

Page 2 **of** 2