PHILIPS
RESPIRONICS

Test Report

Step	Description	Units	Limits	Results	Pass	Fail
176	0050.0180 Device's SD Card Inserted: SD Detected (3) [49m	N/A	EQ YES	YES	✓	
177	0050.0190 Device's SD Card Inserted: SD Recognized (3) [4	N/A	EQ YES	YES	✓	
178	0050.0195 Device's SD Card Inserted: SD Unlocked (3) [49m	N/A	EQ YES	YES	✓	
179	0050.0200 Significant Event Log (190 entries): [52m 6s]	N/A	Pass	TRUE	✓	
180	0050.0210 Engineering Debug Log (124 entries): [54m 11s]	N/A	Pass	TRUE	✓	
181	0050.0020 Reset Operational Hours (102,5 Hrs) [54m 11s]	N/A	Pass	TRUE	✓	

Test Started On 03/27/24 08:20:38

Serial Number TV112010986

Elapsed Test Time 54m 11s

Status **PASS**