Test Software P/N 4102483

Tested Part Number 1040000

Test Software Name Trilogy FSA (REPAIR TEST)



Software Revision 24.0.0.0

Prepared By P. Pascal

Test Report

Step	Description	Units	Limits	Results	Pass Fail
36	0021.0570 O2 Positive Flow Cal: dPl at 64,2 (SP 65 @72,3P)	ADC Cnts	LE 1990	1853	
37	0021.0580 O2 Positive Flow Cal: dP1 at 73,8 (SP 75 @71,6P	ADC Cnts	LE 2124	1991	√
38	0021.0590 O2 Positive Flow Cal: dP1 at 84,7 (SP 85 @70,8P	ADC Cnts	LE 2256	2125	√
39	0021.0600 O2 Positive Flow Cal: dP1 at 94,8 (SP 95 @70,0P)	ADC Cnts	LE 2394	2257	√
40	0021.0610 O2 Positive Flow Cal: dP1 at 105,7 (SP 105 @69,	ADC Cnts	LE 2499	2396	√
41	0021.0620 O2 Positive Flow Cal: dP1 at 115,2 (SP 115 @68,	ADC Cnts	LE 2596	2500	√
42	0021.0630 02 Positive Flow Cal: dPl at 124,6 (SP 125 @67,	ADC Cnts	LE 2704	2597	√
43	0021.0640 O2 Positive Flow Cal: dP1 at 135,2 (SP 135 @66,	ADC Cnts	LE 2794	2704	√
44	0021.0650 O2 Positive Flow Cal: dP1 at 145,2 (SP 145 @65,	ADC Cnts	LE 3026	2795	√
45	0021.0660 O2 Positive Flow Cal: dP1 at 173,6 (SP 175 @61,	ADC Cnts	LE 3157	3027	
46	0021.0670 O2 Positive Flow Cal: dP1 at 188,9 (SP 190 @59,	ADC Cnts	LE 3999	3158	→
47	0021.0880 02 Sensor Cal (NO_O2_ERROR): 592 mV at ambient a	N/A	Pass	TRUE	
48	0021.0930 Write OBM Sensors Cal Table (V: TRUE, A: TRUE)	N/A	Pass	TRUE	

Test Started On 04/25/24 10:08:53

Elapsed Test Time 43m 23s

Status PASS

Serial Number TV018021912

Page 2 **of** 2