Test Software P/N 4102483

Tested Part Number LA1054096

PHILIPS RESPIRONICS

Software Revision 24.0.0.0

Prepared By P. Pascal

## Test Report

Step	Description	Units	Limits	Results	Pass	Fail
36	0020.0570 Pos Flow Cal: dP2 at 77,7 (Setpoint 75) [13m 31	ADC Cnts	LE 48838	47802	√	
37	0020.0580 Pos Flow Cal: dP2 at 87,5 (Setpoint 85) [13m 31	ADC Cnts	LE 49732	48840	✓	
38	0020.0590 Pos Flow Cal: dP2 at 96,8 (Setpoint 95) [13m 31	ADC Cnts	LE 50733	49732	<b>√</b>	
39	0020.0600 Pos Flow Cal: dP2 at 108,0 (Setpoint 105) [13m	ADC Cnts	LE 51554	50734	<b>√</b>	
40	0020.0610 Pos Flow Cal: dP2 at 117,5 (Setpoint 115) [13m	ADC Cnts	LE 52330	51555	<b>√</b>	
41	0020.0620 Pos Flow Cal: dP2 at 127,5 (Setpoint 125) [13m	ADC Cnts	LE 53152	52331	<b>√</b>	
42	0020.0630 Pos Flow Cal: dP2 at 137,7 (Setpoint 135) [13m	ADC Cnts	LE 53811	53154	<b>√</b>	
43	0020.0640 Pos Flow Cal: dP2 at 147,7 (Setpoint 145) [13m	ADC Cnts	LE 55050	53812	<b>√</b>	
4 4	0020.0650 Pos Flow Cal: dP2 at 168,7 (Setpoint 170) [13m	ADC Cnts	LE 56112	55051	<b>√</b>	
45	0020.0660 Pos Flow Cal: dP2 at 187,8 (Setpoint 190) [13m	ADC Cnts	LE 65435	56113	<b>√</b>	
46	0020.0900 Atmospheric Press Sensors Cal Patm: Slope (m) [	lnHg/ADC	EQ 0,000517	0,000517	<b>√</b>	
47	0020.0910 Atmospheric Press Sensors Cal Patm: Intercept (	inHg	1,91 to 4,09	2,57	<b>√</b>	
48	0020.0920 Write Sensors Cal Table [13m 52s]	N/A	Pass	TRUE	<b>√</b>	
49	0020.0970 Write Pprox Cal Table [14m 26s]	N/A	Pass	TRUE	<b>√</b>	
5 0	0020.0935 02 Sensor Cal (not attempted for SW ver.14.2.05	N/A	None	None	<b>√</b>	
51	0020.1030 Write Image Table [14m 47s]	N/A	Pass	TRUE	1	
						ļ
						<del>                                     </del>
					l	<del>                                     </del>
						-
						<del>                                     </del>
						<del>                                     </del>
						<del></del>
						-
						<u> </u>
						<u> </u>

**Test Started On** 04/03/24 11:29:21

Serial Number TV113102307 Elapsed Test Time 14m 47s

Status PASS

**Page** 2 **of** 2